



DU[®] 800 Spectrophotometer Installation and Operating Instructions

512860AD

July 2002

**Beckman Coulter, Inc.
4300 N. Harbor Blvd., Fullerton, CA 92835
© 2002 Beckman Coulter, Inc., Printed in U.S.A.**

Table of Contents

Overview	1
Standard Applications Software	2
Optional Applications Software.....	2
Principles of Operation	4
Optical Principle	4
Blanking.....	4
Reading	5
Scanning.....	6
Read Average Time.....	6
Smoothing	6
Spectrophotometer	7
Computer (PC).....	8
Software	8
Space Requirements.....	9
Power Requirements	10
Unpacking Instrument.....	11
Shipping Kit.....	12
PC, Monitor, and Optional Printer	12
Hardware Installation.....	13
Software Installation	15
General	17
Taking Single Wavelength Readings	18
Editing a Method	19
Blanking.....	22
Reading	23
Print, Save, and Clear.....	24
Taking a Wavelength Scan	25
Editing a Method	26
Blanking.....	28
Scanning.....	28
Print, Save, and Clear.....	30
Performing a Kinetic Run	30
Editing a Method	31
Blanking.....	34

Reading	35
Print, Save, and Clear.....	38
General.....	39
Default Directories.....	39
File Extensions.....	40
Passwords.....	40
Logon Levels	41
On-Line Launch (with System Initialization).....	41
Off-Line Launch	43
Common Front End	43
User Interface.....	43
Main Menu.....	44
Toolbar and Icons.....	45
BLANK, (GO) READ/SCAN, and STOP Buttons.....	45
RediRead and RediScan.....	46
Cell Match.....	46
Edit Method	46
Open File, Save File, and Clear	46
Accessories	48
Method Parameters Frame	48
Status & Control Frame	49
Status Information.....	49
User Status	49
Blank Status	49
UV Lamp Status.....	49
Status & Control Elements	49
Sources.....	49
Holders.....	50
Applications Frame.....	51
Permanent Status Bar.....	51
Operation and Status Message.....	51
Spectrophotometer Status	51
System Menu	52
User Logon / Change User.....	52
Change Password	52
Log Off.....	53
Sources.....	53
Accessories	53
Exit.....	54

Transport/Holder	54
Sipper	55
Batch Sampler	56
Peltier Temperature Controller	59
Manual Control	59
Automatic Control	59
Applications Menu	60
Standard and Optional Applications	60
Custom Applications	61
Methods Menu	61
<Default Method>	61
Select Method	61
Create/Edit Method	62
Delete Current Method	63
Copy Methods from Other Users	63
Copy to Custom Applications	63
Backup/Export Methods	65
Restore/Import Methods	65
Tools Menu	67
RediRead	68
RediScan	70
Function Menu	72
Trace	72
Annotate	72
Zoom	73
Grid	73
Axis Menu	74
Set Scan Limits	74
Autoscale	74
Dynamic Autoscaling	74
Read/Display Mode	75
Cell Matching	75
Data Export	77
Electronic Signatures	78
21 CFR Part 11	78
View Electronic Signature(s)	78
Add Electronic Signature	79
User Options	80
Output Tab	81

Header Tab.....	81
Graph Tab.....	81
Colors Tab.....	81
Administration	81
System Options	82
Setup Tab	82
Output Tab	84
Headers Tab	85
Graph Tab.....	86
Colors Tab.....	87
Regulatory Tab.....	88
User Account Manager	90
Delete Custom Applications	91
View System Audit Trail.....	92
Source Scheduler	93
Performance Validation Scheduler.....	94
Backup/Export System	96
Restore/Import System	96
Add Software Option(s).....	98
Create Backup License File	99
Security/Enhancements.....	100
Lock RediRead and RediScan	100
Force Users to Save Data	101
Data/Methods Security	102
Disallow Creation of Methods by User	102
Show Only Custom Applications	102
Data Directory Structure.....	103
Enhancements	105
Automatically Convert to CSV File.....	105
Diagnostic Menu.....	106
System Status.....	106
User Diagnostics	107
Calibrate Wavelength.....	107
Run Baseline Test	108
Set Scan Gains	108
Reset Source Hours.....	110
Service Diagnostics.....	111
Calibration Info.....	111

Wavelength Drive Run	112
DRP Scan	113
Blocked Beam Noise.....	114
Erase EEPROM.....	114
Help Menu	115
Applications Software.....	116
Transport	117
Cell Holders	118
Sipper Accessory	119
Sipper Controls	119
Batch Sampler Accessory	121
Beckman Coulter (ISCO) Batch Sampler.....	121
Automatic Operation.....	122
Manual Operation	123
Peltier Temperature Controller	123
Standard Controller.....	124
High Performance Controller.....	124
Storage and Transport.....	127
Sample Compartment Configuration	127
Performance Specifications	128
Physical & Environmental Specifications.....	128
General Information.....	129
System Status.....	130
Fuse Replacement	130
UV Source Replacement.....	132
Visible Source Replacement	135
Part I. Replacing the Source.....	135
Part II. Peaking the Source.....	138
Startup Diagnostic Tests	141
Operational Failures.....	142
Operational Messages	143
Warranty Registration	145
Beckman Coulter Offices.....	146
North America	146
United States	146
Canada	146
Latin America	146
Mexico	146

Europe.....	147
France.....	147
Germany.....	147
Italy.....	147
Netherlands.....	147
Spain.....	148
Sweden.....	148
Switzerland.....	148
Turkey.....	148
United Kingdom.....	148
Asia/Pacific.....	149
Australia.....	149
China.....	149
Hong Kong.....	149
Japan.....	149
Singapore.....	150
Taiwan.....	150
Eastern Europe, Middle East, Africa.....	150
South Africa.....	150

List of Figures

Figure 1 DU 800 Spectrophotometer with Computer and Monitor	1
Figure 2 Optical Diagram	4
Figure 3 DU 800 Spectrophotometer	7
Figure 4 DU 800 System and Applications Software	8
Figure 5 Minimum Space Requirements with Desktop PC	9
Figure 6 Minimum Space Requirements with Tower PC	9
Figure 7 Space Requirements with Optional Accessories and Tower PC	10
Figure 8 DU 800 Spectrophotometer on Pallet	11
Figure 9 DU 800 Spectrophotometer Rear Panel Connectors	13
Figure 10 USB Connections on the PC and Spectrophotometer	14
Figure 11 Found New Hardware Wizard Screen	16
Figure 12 User Display Screen	17
Figure 13 Window with Fixed Wavelength Application Selected	19
Figure 14 Method Window for Fixed Wavelength Before the Change	20
Figure 15 Method Window for Fixed Wavelength After the Change	21
Figure 16 Window with Fixed Wavelength Application and all Changes Made	22
Figure 17 Window with Fixed Wavelength Application and Measurement Taken From Three Samples Using a Single Cell Holder	24
Figure 18 Window with Wavelength Scan Application Selected	26
Figure 19 Method Window for Wavelength Scan After the Change	27
Figure 20 Method Parameters Frame After the Change	28
Figure 21 Window with Wavelength Scan Application and Scans Taken From Three Samples Using a Single Cell Holder	29
Figure 22 Window with Kinetics/Time Application Selected	31
Figure 23 Method Window for Kinetics/Time Before the Change	32
Figure 24 Method Window for Kinetics After the Change	33
Figure 25 Window with Kinetics/Time Application and All Changes Made	34
Figure 26 Data Acquisition Mode	36
Figure 27 Rates Mode	37
Figure 28 DU 800 System Initialization Screen	42
Figure 29 Integrated Frame-Oriented User Interface	44
Figure 30 Applications Frame	51
Figure 31 User Logon Screen	52
Figure 32 Change Password for Administrator Screen	53
Figure 33 Accessories Screen with Transport/Holder Tab Selected	54
Figure 34 Accessories Screen with Sipper Tab Selected	55
Figure 35 Accessories Screen with Batch Sampler Tab Selected	56
Figure 36 Accessories Screen with Temp Controller Tab Selected	59
Figure 37 Delete Current Method Screen	63

Figure 38 Copy Methods from Other User Screen	63
Figure 39 Copy Methods to Custom Applications Screen	64
Figure 40 Backup/Export Methods Screen	65
Figure 41 Restore/Import Methods Screen	66
Figure 42 RediRead Screen	69
Figure 43 RediRead Dialog Screen	70
Figure 44 RediScan Screen	71
Figure 45 Scan Without a Blank Dialog	72
Figure 46 Set Scan Limits Screen	74
Figure 47 Cell Matching Screen	75
Figure 48 Example of Data Exported as a CSV File	77
Figure 49 Electronic Signature Authentication Screen	79
Figure 50 Add Electronic Signature Screen	80
Figure 51 System Options Screen with Setup Tab Selected	82
Figure 52 User Authentication Screen	83
Figure 53 User Options Screen with Output Tab Selected	84
Figure 54 User Options Screen with Headers Tab Selected	85
Figure 55 User Options Screen with Graph Tab Selected	86
Figure 56 User Options Screen with Colors Tab Selected	87
Figure 57 System Options Screen with Regulatory Tab Selected	88
Figure 58 Password Renewal Screen	89
Figure 59 User Account Manager Screen	90
Figure 60 Delete Custom Applications Screen	91
Figure 61 System Audit Trail Screen	92
Figure 62 Source Scheduler Screen	93
Figure 63 Performance Validation Scheduler Screen	94
Figure 64 Performance Validation Schedules Activation Dialog	95
Figure 65 Performance Validation Scheduler Screen	95
Figure 66 Performance Validation Scheduler Screen with Error Message	96
Figure 67 Restore/Import System Screen	97
Figure 68 Add Software Options(s) Screen	98
Figure 69 Select License File Screen	98
Figure 70 Add Software Options(s) Screen	99
Figure 71 Administration Menu.	100
Figure 72 21 CFR Part 11 Compliance.	101
Figure 73 Sava Data.	101
Figure 74 Data/Methods Security.	102
Figure 75 Data Directory Structure.	103
Figure 76 Data Directory Structure.	104
Figure 77 Headers for current run.	105
Figure 78 Enhancements	105
Figure 79 System Status Screen	107
Figure 80 Wavelength Calibration Screen	107
Figure 81 Baseline Test Screen	108
Figure 82 Set Scan Gains Screen	109
Figure 83 Reset Source Hours Screen	110

Figure 84 Service Diagnostics Screens	111
Figure 85 Calibration Information Screen	111
Figure 86 Wavelength Drive Run Screen	112
Figure 87 DRP Scan Screen	113
Figure 88 Blocked Beam Noise Screen	114
Figure 89 Erase EEPROM Attention Message Screen	114
Figure 90 About This Program Screen	116
Figure 91 Transport Accessory	117
Figure 92 Cell Holders	118
Figure 93 Sipper Accessory	119
Figure 94 Batch Sampler Accessroy	121
Figure 95 Peltier Temperature Controller	124
Figure 96 Diagram of Location and Size of the Beam in Sample Compartment	127
Figure 97 System Status Window Screen	130
Figure 98 Fuse Compartment	131
Figure 99 Location of Source Access Door on DU 800	133
Figure 100 Unplugging UV Source Connector	133
Figure 101 Unscrewing Thumb Screws securing UV Source	134
Figure 102 Removing UV Source	134
Figure 103 Reset Source Hours Screen	135
Figure 104 Location of Source Access Door on DU 800	136
Figure 105 Unscrewing Thumbscrews Holding Visible Source	137
Figure 106 Removing Visible Source	137
Figure 107 Installing New Visible Source	138
Figure 108 Using Peaking Tool	139
Figure 109 Reset Source Hours Screen	140
Figure 110 DU 800 System Initialization Screen	141



Introduction


Overview

The DU[®] 800 Spectrophotometer is a PC controlled system intended for use in quantitative and qualitative analysis in biological and industrial procedures that require spectrophotometric measurements in the UV and visible region of the electromagnetic spectrum. If the instrument is used in a manner other than as described, the safety and performance of the instrument can be impaired.

The DU 800 Spectrophotometer operates in the wavelength range of 190 to 1100 nm and has a bandwidth of ≤ 1.8 nm. The focused micro-beam design provides a wide linear range and other specific benefits for small volumes and precious samples. Various accessories are available to address micro-volume samples and individual application requirements.

The control of the instrument, data handling, and data reduction capabilities are contained within the confines of the PC. The frame-oriented software provides a convenient and user-friendly interface and, therefore, ensures a quick learning curve. The PC must run Windows 2000 as an operating system. The general operation of the software is described in System Software and the standard and optional applications in Applications Software.

Figure 1 DU 800 Spectrophotometer with Computer and Monitor



Standard Applications Software

The following standard applications are available after software installation:

Fixed Wavelength - Performs Absorbance or %Transmittance readings for up to 12 wavelengths simultaneously. A factor or custom formula may be applied to calculate final results.

Wavelength Scan - Performs wavelength scans in Absorbance or %Transmittance. Acquired scan data are stored and may be used for various manipulations and calculations. Includes the calculation mode to add, subtract, multiply and divide spectra.

Kinetics/Time - Simultaneously measures and analyses up to 12 rate reactions. Data can be reviewed in real time and automatically calculated and printed. The rate of the kinetic reaction is reported using a linear regression.

Nucleic Acid Analysis - Determines protein impurity in nucleic acid samples based upon the ratio of readings at two wavelengths with a choice of background correction. Protein and nucleic acid concentrations can also be calculated using the Warburg and Christian¹ coefficients.

Single Component Analysis - Determines the concentration of unknowns by either linear or non-linear (quadratic) regression and provides statistical analysis for the standard curve. The standard curve can have up to 30 standards. The operator can re-run any standard, remove or add standards, and have the instrument re-calculate the curve.

Performance Validation - Provides a simple procedure to verify the performance of the instrument without standards or samples. Tests include: wavelength accuracy and repeatability, resolution, baseline flatness, noise, and stability.

In addition, the "rapid" modes **RediRead** and **RediScan** are available. These provide the user with a fast and easy way to take readings at fixed wavelengths or to make wavelength scans.

A detailed description of the above applications can be found in Applications Software.

Other features, such as *User Logon*, *User Customization*, *Methods Backup and Transfer*, *Diagnostics*, *Source Scheduler*, and *Performance Validation Scheduler* are described in System Software.

Optional Applications Software

Wavelength Scan II - Extends the functionality of Wavelength Scan with the following modes: Derivatives (1st, 2nd, 3rd, and 4th), NetA Calculations, Scatter Correction, Point Picker, and Peak/Valley Picker. Also included is the display of spectra in LogA.

Nucleic Acid Analysis II - Extends the functionality of Nucleic Acid Analysis with multiple ratios and other concentration calculations. Also included are DNA/RNA Oligo Quantitation modes to determine molecular weight, absorptivity (extinction

coefficient), concentration, and the theoretical melting point for oligonucleotide DNA samples.

Protein Analysis - Calculates protein concentrations using the Bradford, Lowry, Biuret, Direct UV, Colloidal Gold, or Bicinchoninate (BCA) methods. The user may choose to add, delete, or re-run individual standards based upon the statistical analysis of the standard curve.

Enzyme Mechanism - Provides for rapid and easy characterization of a wide variety of enzyme reactions. The software calculates and reports K_m , V_{max} , k_{cat} and K_i as well as the Hill constant. The following functions are available: Michaelis-Menten, Lineweaver-Burk, Eadie-Hofstee, Hanes-Woolf, Hill, and Inhibitor plots.

Enzyme Activity - Calculates the enzyme activity of large numbers of samples from chromatography fractions.

Experimental T_m Analysis - Allows the study of denaturation and renaturation of DNA samples. The thermal melting point (T_m) is determined quickly and efficiently using the First Derivative, 2-Point Average, or a Non-Linear Curve Fit algorithm. Up to six micro-volume samples with 325 μ L can be processed automatically, applying up to three temperature ramps.

A full line of modular accessories is available: ambient and temperature-controlled single and multi-position cell holders, sipper accessory, and batch sampler.

Accessories to support micro-volume sampling include microcell holders for up to 12 samples, the 50 μ L Microcell, the 100 μ L Multi-Microcell, and the 5 μ L Ultra-Microcell.

A detailed description of the above applications can be found in Applications Software.


¹ Warburg, O. and Christian, W., Biochem Z. p. 384f (1942).

Principles of Operation

Optical Principle

The DU 800 Spectrophotometer is a single beam instrument. Light from both sources enters the monochromator where it is dispersed by a concave holographic grating. Monochromatic light exits the monochromator and illuminates the sample. The amount of light that passes through the sample is measured by a single photodiode detector.

Figure 2 Optical Diagram



The focal point of the beam in the sample compartment is on the right-hand side. All sampling accessories position the sample at the focal point for best performance with regular samples and micro-samples.

Blanking

A blank is always required before data collection; any reading without a blank is invalid. A blank reading is taken when the BLANK icon button is clicked. This icon button is located in the toolbar.

NOTICE

In the RediRead and the RediScan modes, the blank is executed in the Commands menu or by clicking on the Blank button.

When the instrument blanks, the following steps are performed:

1. The monochromator is moved to the proper wavelength. This is the specified wavelength for a single wavelength reading.
2. The proper detector gain value is selected automatically. This minimizes the noise level and maximizes photometric accuracy.
3. Dark current is measured and corrected. This compensation assures accurate readings at high absorbance.
4. In the Wavelength Scan application only, a background scan is taken. The blank (or reference) is automatically scanned over the same range at the same speed that the sample will be scanned, so that the background correction is optimal.

This calibration assures repeatable readings every time the instrument is used.

In all modes, a blank solution should be in the sample compartment during the blank. It is suggested that the solvent used to prepare the samples be used for the blank. However, air (no sample) may be used. A new blank reading should be taken each time the solvent is changed.

NOTICE

Plastic cuvettes, glass (Pyrex) cuvettes, and some solvents have significant absorption in the UV region. Verify that they transmit UV light by scanning them versus air before using them in the UV region.

To re-zero the instrument at any time between samples, insert the same blank solution and click on the BLANK icon button.

The instrument stores the blank and uses it until either the sources are turned off, another blank reading is taken, or certain parameters are changed. For best results, the instrument should be blanked frequently, allowing the blank reading to be taken shortly before the sample measurement is taken. A new blank should be read if the instrument has not been used for an hour.

Reading

The background reading, taken as part of the Blank procedure, is stored in the instrument and can be reused for an unlimited number of sample readings as long as the wavelengths remain the same. When a new blank is required, it is indicated in the Status & Control Frame.

A new blank reading should be taken every time a solvent is changed because the blank reading will likely be different. A new blank reading should also be made if no reading has been taken for over an hour. To blank, click on the BLANK icon button while in the appropriate application.

For multiple wavelengths (up to 12), a blank reading is taken and stored for each selected wavelength.

As the sample data are collected, the blank is subtracted and the difference in absorbance (or transmittance) is reported.

Scanning

The background scan, taken as part of the Blank procedure, is stored in the instrument and can be reused for an unlimited number of sample scans as long as the range and scan speed remain the same. (The range can be decreased as long as the scan speed remains constant and no blank is required.) When a new blank is required, it is indicated in the Status & Control Frame.

A new blank scan should be made every time a solvent is changed because the blank spectrum will likely be different. A new blank scan should also be made if no scan has been taken for over an hour. To re-scan the blank, click on the BLANK icon button while in Wavelength Scan.

The selected scanning speed determines the distance between each data point that is collected as the instrument scans through the chosen region. At 1200 nm/min, a data point is collected every nanometer. At 600 nm/min, a data point is collected every half nanometer.

As the sample data are collected, the blank is subtracted and the difference in absorbance (or transmittance) is reported.

Read Average Time

The noise level of the instrument, and therefore the uncertainty of a sample reading, is decreased by taking a number of readings and averaging them. The instrument takes a reading every 0.05 second. It takes a series of these readings over a user-selected time and averages them to obtain the blank and sample readings. For example, with a read average time of 0.5 seconds, ten readings are taken and averaged. The operator can specify a read average time from 0.05 to 99.9 seconds in all applications, except *Wavelength Scan* and *RediScan*.

Background and sample scans are collected without averaging. Smoothing may be used to improve the appearance of scan data.

Smoothing

A wavelength scan can be smoothed using a selectable smoothing function. The calculation, using the Savitzky and Golay¹ coefficients (as modified for end points by Peter A. Gorry²), is done for every data point in the scan, using the data points before and after the point of interest. The software selects the degree of smoothing, based on the user selection.

Medium smoothing uses a calculated amount of data points and causes a pre-defined degree of smoothing. Light smoothing uses half the calculated data points and, therefore, causes a lighter degree of smoothing. Heavy smoothing uses twice the calculated data points and, therefore, causes a higher degree of smoothing.

¹ Savitzky, A., and Golay, M., *Anal Chem*, 36, 1964, p1627f.

² Gorry, Peter A., *Anal Chem*, 62, 1990, p570f.

Installation

This chapter describes how to install the DU Series 800 Spectrophotometer and the required personal computer. Installation by a qualified Beckman Coulter Field Service Engineer can be purchased by contacting your local Beckman Coulter sales office.

The DU 800 UV/Visible Spectrophotometer is designed to sit on a laboratory bench or table, which is level, flat and capable of supporting its weight and the weight of all accessories.


The instrument is designed to operate in a clean laboratory environment, free from dust, fumes, excessive moisture, and corrosive chemicals. It should not be exposed to drafts from heating or cooling vents, heating elements, open windows or doors. Lab areas that receive direct sunlight should also be avoided.

An ambient temperature of 15-40°C (59-104°F) should be maintained. Relative humidity should be 85% or less.

Instrument performance can be affected by strong electromagnetic fields that can exist in the proximity of large electric motors, centrifuges, diathermy machines, and microwave sources.

Spectrophotometer


Figure 3 DU 800 Spectrophotometer



The DU 800 includes:

- DU 800 UV/Visible Spectrophotometer (optical bench)
- System and Applications Software CD-ROM
- USB Cable, 6-foot
- PC with pre-installed software, *if ordered*.

Computer (PC)



A computer (PC) is required to operate the DU 800 UV/Vis Spectrophotometer.

Minimum Requirements:

- Standard PC (266MHz or faster)
- Available USB Port
- 128MB RAM
- 3.5-inch Floppy Drive
- CD-ROM Drive
- Microsoft Windows 2000

Beckman Coulter fully supports tested PC models, which includes the optional PC that can be ordered from Beckman Coulter. Other PC's that meet the minimum requirements and have the appropriate operating system installed, should be compatible. However, this cannot be guaranteed by Beckman Coulter.

The optional PC from Beckman Coulter ensures a ready-to-use system and comes in a fixed configuration, which is subject to change:

- IBM PC with USB Interface
- 15-inch Monitor or Flat-Panel Display
- Microsoft Windows 2000 or later (pre-installed)
- DU 800 System and Applications Software (pre-installed)

The pre-installed software is included as an image on CD-ROM(s), which allows the user to restore the entire content of the hard disk.

Software

Figure 4 DU 800 System and Applications Software



The DU 800 System and Applications Software has been validated with the following local-language Windows 2000 Operating Systems:

- U.S. English
- German
- Japanese

The software was written following Microsoft Windows development guidelines and, therefore, should be fully compatible with other Latin and Asian languages.

Follow the instructions found in paragraphs Unpacking, Hardware Installation and Software Installation to install the system.

Space Requirements

The PC must be located close enough to the spectrophotometer so that the interconnecting USB cable will reach the USB port on the instrument. If desired, the monitor can be placed on the top left side of the DU 800 spectrophotometer. Large monitors may restrict the movement of the sample compartment cover.

If purchased, the batch sampler should be placed adjacent to the right hand side of the instrument so to minimize the length of tubing needed to reach the sample compartment. The other accessories can be placed in a convenient location near the instrument, within reach of the interconnecting cables.

The spatial requirements for the DU 800, PC, and accessories are diagrammed in the following Figures. Dimensions shown in cm and (inches). The shaded areas show additional space required for air circulation. Do not block these air spaces.

Figure 5 Minimum Space Requirements with Desktop PC






Figure 6 Minimum Space Requirements with Tower PC



The following Figure shows a configuration with other optional accessories, such as Peltier Temperature Controller, Sipper Accessory, and Batch Sampler.

Figure 7 Space Requirements with Optional Accessories and Tower PC



901217L.A1

Power Requirements

CAUTION

The DU 800 Spectrophotometer, PC, and monitor must be plugged into grounded electrical outlets.

If the plug on the power cord is not compatible with the electrical outlet, contact the local Beckman Coulter office. Do not use the power cord in this case.

High leakage current. Ensure proper grounding.

The following optional accessories also require a grounded electrical outlet: Printer, Batch Sampler, and Peltier Temperature Controller. The electrical requirements for the DU 800 Spectrophotometer, Batch Sampler, and Peltier Temperature Controller are listed in the following Table. The electrical requirements for the PC, monitor, and printer can be found in the respective installation manual of each item.

Table 1 Electrical Requirements

	Frequency (Hz)	Voltage (VAC)	Current (Amps)
Spectrophotometer	50/60	100-240V±10%	3.0
Batch Sampler (automatically detects proper voltage)	50/60	100V-120V±10%/220V-240V ±10%	1.5 0.8
Peltier Temperature Controller	50/60	100V-120V ±10%/220V-240V ±10%	1.0 0.5

The number of boxes received will depend on the items ordered. Minimally you will receive one box, which contains the DU 800 Spectrophotometer and the shipping kit as well as regular accessories. If purchased, you may also receive a personal computer and the monitor in the box. Additionally ordered accessories may be shipped in separate boxes.


Once unpacked, inspect all parts for shipping damage and, if necessary, contact the carrier regarding any damage caused in shipment as soon as possible. Note that the carrier is responsible for damage occurred while in transit.

Use the respective packing slips to verify that the listed items of each package were received. Report any shortages to the local Beckman Coulter sales office.

Unpacking Instrument

The original shipping container for the DU 800 Spectrophotometer is attached to a pallet. If not already removed from the pallet, cut the straps that hold the shipping container together.

Figure 8 DU 800 Spectrophotometer on Pallet



Remove the straps and the instrument box, which sits on the pallet on top of the instrument (and the PC, if purchased). Locate the *Quick Installation Sheet* and any other documentation that may be present and put it aside.

The *Quick Installation Sheet* can be used to quickly setup the DU 800 Spectrophotometer in a step-by-step fashion. More information on installation is provided in the *DU 800 Installation and Operating Instructions*.

NOTICE

Complete and return the Warranty Registration after installation.

This will guarantee that your system will be registered and that the one-year warranty is applied properly. It also assures quick and easy access to the Beckman Coulter Product Support team for answers to your questions.

Please remove the box located on the right side of the instrument. This box is the *Shipping Kit*.

Remove any packing material that covers the instrument and carefully place the DU 800 at the chosen location.

Shipping Kit

Please locate the DU 800 Shipping Kit, which is packed in the large DU 800 shipping box. The Shipping Kit contains the various items required for the operation of the DU 800 Spectrophotometer. Verify that the contents of the shipping kit match the items shown on the included list.

PC, Monitor, and Optional Printer

The DU 800 Spectrophotometer is controlled by the *DU 800 System and Applications Software* running under Microsoft Windows 2000. The required PC and monitor and/or the optional printer may have been purchased from Beckman Coulter, or they may have been acquired from another source.


If the PC was not purchased from Beckman Coulter, please verify that it meets the minimum requirements.

Please refer to the manuals included with the PC, monitor and printer for unpacking information.

Hardware Installation

1. Place the DU 800 Spectrophotometer in the chosen location, and allow access to the back panel until installation is complete.

Figure 9 DU 800 Spectrophotometer Rear Panel Connectors



2. Attach the power cord compatible with your country's voltage outlet to the back of the DU 800 and plug the other end into a grounded electrical outlet. The instrument contains a switching power supply that covers the range from 100 - 240 VAC.


NOTICE

Do not power on the DU 800 at this time

3. If purchased, the installation of the Sipper Accessory, Batch Sampler and/or Peltier Temperature Controller should be performed now. Follow the installation instructions supplied with each accessory.

4. Place the PC in a convenient location near the DU 800 and attach the flat connector of the USB cable to the USB port of the PC. Attach the other end of the connector to the USB port located on the back of the DU 800.

Figure 10 USB Connections on the PC and Spectrophotometer



NOTICE

It is recommended that you label the USB Port on the PC as "DU 800" and always use this port for the connection to the Spectrophotometer. Switching the cable to a different port after the DU 800 software is installed will require a manual installation of the drivers.

5. Setup the PC with monitor and peripherals. Make sure that it meets the minimum requirements (see Requirements, Computer). Refer to the installation instructions included with the PC, monitor and printer for information on how to setup and power on each item.

NOTICE

The monitor can be placed on the top left side of the DU 800.

Large monitors may restrict the movement of the sample compartment cover.

Software Installation

1. Power on the PC.
2. Locate the *DU 800 System and Applications Software* CD provided in the shipping kit and place it in the CD-ROM drive of the PC.
3. The setup should start automatically after inserting the CD. If it does not, access the CD-ROM drive from Windows Explorer and click on SETUP.EXE to start the installation manually.
4. Follow the on-screen instructions to complete the software installation.
5. Power on the DU 800. The power switch is located on back of the DU 800 Spectrophotometer. The green Power LED on the front of the DU 800 will illuminate.

NOTICE

If the green Power LED does not illuminate, turn off the instrument and check the fuse. Directions for fuse replacement are provided in Maintenance, Fuse Replacement.

See Troubleshooting, Operational Failures, for a complete list of problem resolutions.

6. The PC will automatically detect the DU 800 Spectrophotometer connected via the USB port.
7. When a DU 800 is plugged into the USB port, Windows will try to identify the device and install the appropriate drivers automatically. When following the described installation procedure, the required USB drivers will be installed automatically and the installation will complete. In this case, skip the following and go to Step 8.

If the *DU 800 System and Applications Software* has not been installed yet or the drivers cannot be found for another reason, a window similar to the one shown below will appear and the required drivers, which are located on the *DU 800 System and Applications Software* CD-ROM, may be added manually. However, we highly recommend that you cancel the *Found New Hardware Wizard* window and install the *DU 800 System and Applications Software* before turning on the instrument. If manual installation is necessary, the DU 800 drivers can be found on the System and Applications Software CD-ROM in the \Window\inf and the \Window\System 32\Drivers directories.

Figure 11 Found New Hardware Wizard Screen



NOTICE

Refer to Troubleshooting, Operational Failures if you encounter any problems during installation

8. You may now launch the software and operate the DU 800 Spectrophotometer (see Launch and Initialization).

Getting Started

General

Make sure that you have reviewed the section Introduction before you begin to explore the system and its operation.


In order to make measurements, the DU 800 System and Applications Software must be running and the DU 800 Spectrophotometer must be initialized. The section Launch and Initialization describes the necessary steps.

The following step-by-step instructions will lead you through the basic applications, using simple examples. Please refer to Applications Software for more information.

Install a cell holder in the sample compartment of the instrument. Configure the software for this particular holder in the Transport/Holder tab of the *Accessories* window. For the following examples, a *Single Cell Holder* should be installed and configured. With a *Multicell Holder* installed and configured, up to 12 samples can be automatically processed within a set. However, this is not described in the following step-by-step exercises. In this case, it is recommended to set the number of cells in the Sampler Tab of the Method of the selected application to 1, before starting a sample run, and use a single cell position.

By default, the system uses the <Generic User> level.

Figure 12 User Display Screen



The system includes a Visible Lamp (Tungsten) and a UV Lamp (Deuterium-Halogen). The visible lamp covers the range from 321 to 1100 nm while the UV lamp provides the energy for the wavelength range from 190 to 415 nm.

For the following examples, both lamps should be turned on. Click on the *Visible* icon button to turn on the visible lamp and then click on the *UV* icon button.



The visible lamp is turned on immediately while the UV lamp requires approximately 30 seconds warm up time before it can be used. The warm up period is indicated by the status message "Warming up UV lamp ..." while the UV icon button flashes.

When a lamp is turned on, the respective menu item is checked and the caption of the respective icon button turns red.

After the UV lamp has been warmed up the message in the Status & Control Frame reads "UV Lamp On for x sec" (the unit can be seconds, minutes, or hours). This provides additional information for the user in regard to the UV lamp warm up time.

This sections provides step-by-step operating instructions for three simple standard applications.

Taking Single Wavelength Readings

Taking a Wavelength Scan


Performing a Kinetic Run

Taking Single Wavelength Readings

All applications, except Wavelength Scan I and II, take one or multiple single wavelength readings in one form or another (e.g., reading at 260.0 nm and 280.0 nm).

Make sure that the **Fixed Wavelength** application is selected and that the current method is the **<Default Method>** by checking the drop-down list boxes in the Toolbar. If not, click on the down arrow of the drop-down list box and select the appropriate item. You should see a window similar to the one shown below.

Figure 13 Window with Fixed Wavelength Application Selected



You are now ready to take single wavelength readings at three wavelengths, 200.0, 250.0, and 300.0 nm. But HALT - before we take the first reading we want to change some parameters, which is done in the Method window.


Editing a Method

The <Default Method> for Fixed Wavelength provides a set of pre-defined parameters. For example, there are 3 wavelengths which are specified as 200.0, 250.0 and 300.0 nm.




We want to change these parameter to fit our requirements. Click on the *Create/Edit Method* menu item or the *Edit Method* icon button. This brings up the window shown below with the default parameters for the Fixed Wavelength application.

Figure 14 Method Window for Fixed Wavelength Before the Change



We want to change the number of wavelengths to 1 and the wavelength to 260.0 nm. First, select 1 from the Number of Wavelengths drop-down list box. You will see that the grid with the wavelengths definition is reduced to one (1) wavelength. Now, click on the cell in row 1, which shows 200.0 as the wavelength, change the value to 260 and press Enter. The wavelength is now set to 260.0 nm and your method window should look like the one shown below.

Figure 15 Method Window for Fixed Wavelength After the Change




In this example, we will use the modified default parameters temporarily. To do so, we confirm the changes by clicking on the OK button. In this case, the selected parameters remain only valid until you leave the Fixed Wavelength application or select a new method.

You may save the modified default method parameters permanently under an appropriate method name by clicking on the Save As button. In this case, the next time you use the Fixed Wavelength application, just select your saved method from the Method menu or the method drop-down list box and you are ready-to-go using your individual parameters. You may change your method parameters at a later time or save the method under a different name.

We have now set up the Fixed Wavelength application in a way to fit our needs by modifying the method, in this case the <Default Method>, and your window should now look like the one shown below.

Figure 16 Window with Fixed Wavelength Application and all Changes Made



The following three things have changed when we closed the Method window. First, the Method Parameter List has been updated. Notice that it only shows one wavelength with the entered wavelength value. Second, the data grid changed to only one column with the correct heading to reflect the current parameter settings. And third, the large-number display has only the field that is labeled with the respective unit and wavelength.

We are now ready to measure our samples using customized parameters. But before we take a reading of the first sample, we need to blank the system on the selected wavelength.

Blanking

A blank is required when lamps are turned on or certain parameters change (see Blanking Method for more information). In these cases, the instrument will remind the user with the message BLANK REQUIRED. Insert an empty cuvette or a cuvette with a blank solution in the cell holder at the measurement position and close the sample compartment.



Now click the BLANK button. The *Status & Control Frame* displays Blanking... during the blanking and Last blanked at 11:13 AM when it is complete. In this example, a single blank reading is taken at 260.0 nm.

The system is now ready to provide accurate readings.

Reading

Insert a cuvette containing a sample into the Single Cell Holder (or the appropriate position(s) of the Multi-Position Cell Holder) and close the sample compartment.


The *Sample ID* for the current sample (Sample 1-1) is a default and is shown in blue. With a Multi-Position Cell Holder, there would be a set of default Sample ID's (e.g., from Sample 1-1 to Sample 1-6). You can modify those *Sample ID's* that are shown in blue by clicking on the respective cell and changing the text, followed by the Enter key. This must be done before the measurement of the sample(s) and allows you to use customized *Sample ID's* for the entire set. After a sample reading is complete, the respective *Sample ID* turns black and the cell is locked.



Now click on the READ button to take a reading. After the reading has been taken, the result is reported in the data grid and the system is now ready to process the next sample (or sample set). Open the sample compartment, replace the cell with the next sample (or sample set when using a multi-position cell holder), and then close the sample compartment to read the next sample.


Your window should now look like the one shown below.

Figure 17 Window with Fixed Wavelength Application and Measurement Taken From Three Samples Using a Single Cell Holder



After we have taken the readings of all samples, we must decide what to do with the acquired data. We can discard them, print them, and/or save them. In any case, to complete the current run, select another application or method, or proceed with something else, we must execute *Save* and/or *Clear*.

Print, Save, and Clear



Print - To print the acquired data, select *Print* from the File menu in the Applications Frame.

 Save File

Save - To save the acquired data, select *Save As* from the File menu in the Applications Frame or click on the *Save File* icon button.

The *Save File* dialog appears, which lets you enter an appropriate filename. Click on the OK button to save the data or Cancel to abort. If you don't change the target directory, the data file (.DUX) is saved to the following location "C:\DU800 Data\Generic User\Fixed Wavelength\

 Clear

Clear - To end the current task and reset the application (e.g.; for another task or to leave the application), select *Clear* from the File menu in the Applications Frame or click on the *Clear* icon button. A message will inform you if the data has not been saved. The data grid and the graph are cleared, any unsaved data is discarded, and the application is reset.


You may then start another *Fixed Wavelength* task or select another application.

Taking a Wavelength Scan

In comparison to all other applications, the Wavelength Scan application (I or II) takes scans, which are represented by a continuous wavelength range (e.g., readings from 200.0 nm to 800.0 nm in 1.0 nm intervals).

Make sure that the **Wavelength Scan** application is selected and that the current method is the **<Default Method>** by checking the drop-down list boxes in the Toolbar. If not, click on the down arrow of the drop-down list box and select the appropriate item. You should see a window similar to the one shown below.

Figure 18 Window with Wavelength Scan Application Selected



You are now ready to take wavelength scans from 200.0 to 800.0 nm at 1200 nm/min. But HALT - before we take the first reading we want to change some parameters, which is done in the Method window.

Editing a Method

The <Default Method> for Wavelength Scan provides a set of pre-defined parameters. For example, the wavelength range is specified from 200.0 to 800.0 nm and the scan speed at 1200 nm/min.




We want to change these parameter to fit our requirements. Click on the *Create/Edit Method* menu item or the *Edit Method* icon button. This brings up the Method window with the default parameters for the Wavelength Scan application.

We want to change the Start Wavelength to 500.0 nm, leave the End Wavelength at 800.0 nm, and then set the Scan Speed to 2400 nm/min. This new scan speed allows us to scan faster but at the expense of the scan resolution. Instead of a 1.0 nm interval at 1200 nm/min, the data acquisition for a scan at a speed of 2400 nm/min is performed with a 2.0 nm interval.

First, change the Start Wavelength to 500.0 nm. Then change the scan speed to 2400 nm/min. Select 2400 from the Scan Speed drop-down list box. The Interval value will be recalculated and should now show 2.0 nm as in the following window.

Figure 19 Method Window for Wavelength Scan After the Change



In this example, we will use the modified default parameters temporarily. To do so, we confirm the changes by clicking on the OK button. In this case, the selected parameters remain only valid until you leave the Wavelength Scan application or select a new method.

You may save the modified default method parameters permanently under an appropriate method name by clicking on the Save As button. In this case, the next time you use the Wavelength Scan application, just select your saved method from the Method menu or the method drop-down list box and you are ready-to-go using your individual parameters. You may change your method parameters at a later time or save the method under a different name.

We have now set up the Wavelength Scan application in a way to fit our needs by modifying the method, in this case the <Default Method>. The current parameter settings are reflected in the Method Parameters Frame.

Figure 20 Method Parameters Frame After the Change

Method Parameter	Value
Sampling Device	Holder
Holder Type	Single Cell
Number of Cells	1
Scan Mode	Abs
Start Wavelength	500.0 nm
End Wavelength	800.0 nm
Wavelength Interval	2.0 nm
Scan Speed	2400 nm/min
Auto Print	No
Auto Save	No
Filename Prefix	Default

The Method Parameters Frame has been updated when we closed the Method window. Notice that it now shows the Start Wavelength with 500.0 nm, the Wavelength Interval with 2.0 nm, and the Scan Speed with 2400 nm/min.

We are now ready to scan our samples using customized parameters. But before we take a scan of the first sample, we need to blank the system with the selected wavelength range.

Blanking

A blank is required when lamps are turned on or certain parameters change (see Blanking Method for more information). In these cases, the instrument will remind the user with the message BLANK REQUIRED. Insert an empty cuvette or a cuvette with a blank solution in the cell holder at the measurement position and close the sample compartment.



Now click the BLANK icon button. The *Status & Control Frame* displays Blanking... when a blank is taken and Last blanked at 11:13 AM when it is completed. In this example, a blank scan is taken from 500.0 nm to 800.0 nm with an interval of 2.0 nm at a scan speed of 2400 nm/min.

The system is now ready to provide accurate scans.

Scanning

Insert a sample into the Single Cell Holder (or the appropriate position(s) of the Multi-Position Cell Holder) and close the sample compartment.

The grid below the scan window will be automatically populated when we take scans. The Use check box allows you to select if the scan should be used and displayed or not. By default, each acquired scan will be used and displayed.




Now click on the SCAN button to take a scan. You can see how the readings for each wavelength within the range are taken in real time. If the Dynamic Autoscaling item in the Axis menu is checked, the y-axis will be automatically resized during the scanning process. If not, you can select the Autoscale Y item in the Axis menu to autoscale the y-axis manually when the scan is finished.

You may now open the sample compartment, replace the cell with the next sample (or sample set when using a multi-position cell holder), and then close the sample compartment to read the next sample.


Your window should now look like the one shown below.

Figure 21 Window with Wavelength Scan Application and Scans Taken From Three Samples Using a Single Cell Holder



After we have taken the scans of all samples, we must decide what to do with the acquired scans. We can discard them, print them, and/or save them. In any case, to complete the current run, select another application or method, or proceed with something else, we must execute *Save* and/or *Clear*.

Print, Save, and Clear



Print - To print the acquired data, select *Print* from the File menu in the Applications Frame.



Save - To save the acquired data, select *Save As* from the File menu in the Applications Frame or click on the *Save File* icon button.

The *Save File* dialog appears, which lets you enter an appropriate filename. Click on the OK button to save the data or Cancel to abort. If you don't change the target directory, the data file (.DUX) is saved to the following location "C:\DU800 Data\Generic User\Wavelength Scan\<<Filename>.dux", if the administrator has not changed the default DATA root directory previously.



Clear - To end the current task and reset the application (e.g.; for another task or to leave the application), select *Clear* from the File menu in the Applications Frame or click on the *Clear* icon button. A message will inform you if the data has not been saved. The data grid and the graph are cleared, any unsaved data is discarded, and the application is reset.

You may then start another *Wavelength Scan* task or select another application.


Performing a Kinetic Run

All applications, except *Wavelength Scan*, take single wavelength readings in one form or another. A *Kinetics Run* observes the absorbance of a single wavelength from a sample and uses the change of the absorbance over time to calculate a rate, usually expressed as dAbs/min or dA/min.

Make sure that the **Kinetics/Time** application is selected and that the current method is the **<Default Method>** by checking the drop-down list boxes in the Toolbar. If not,

click on the down arrow of the drop-down list box and select the appropriate item. You should see a window similar to the one shown below.

Figure 22 Window with Kinetics/Time Application Selected



You are now ready to start the Kinetic run. But HALT - before we take the first reading we want to change some parameters, which is done in the Method window.


Editing a Method

The <Default Method> for Kinetics/Time provides a set of pre-defined parameters. For example, in the <Default Method>, the analytical wavelength is set to 540.0 nm and samples are supposed to be observed over a total time of 120 seconds, taking a measurement for each sample every 15 seconds. In addition, we want to monitor the rate during the run to see if the absorbance change over time is within the expected range.



We want to change these parameter to fit our requirements. Click on the *Create/Edit Method* menu item or the *Edit Method* icon button. This brings up the window shown below with the default parameters for the Kinetics/Time application.

Figure 23 Method Window for Kinetics/Time Before the Change



We want to change the number of samples to 1 and correct the measurement for background at 480.0 nm.

First, click on the Background Wavelength field and enter 480. Then check the Background Correction box to determine that a measurement will also be taken at the background wavelength of 480.0 nm, in addition to the measurement at the analytical wavelengths. The result of the two measurements will be an absorbance corrected for background, a so called Net Absorbance (absorbance - background absorbance).

Second, click on the Samples field in the Kinetic frame and enter 1. This determines that the Kinetic run will consist of a single sample.

Third, change the Interval Time and Total Time to values that better suit this experiment. Set the Interval Time to 5 seconds and the Total Time to 60 seconds.

Finally, check the Delta Abs box so we can monitor the rate (dAbs/min) for each time point, starting with the second. This absorbance change at any time point gives us an indication about the time course of the kinetics and would allow us to make a decision in order to abort the kinetic run, if the rate of the absorbance change is not what we expect it to be.

All parameters have been set and your method window should look like the one shown below.

Figure 24 Method Window for Kinetics After the Change

Method for Kinetics/Time

<Default Method>

Application | Sampler | Report | Info

Wavelengths:

Analytical Wavelength: 540.0 nm

Background Wavelength: 480.0 nm

Background Correction

Kinetic:

Samples: 1

Interval Time: 5.0 Seconds

Total Time: 60.0 Seconds

Factor / Units: 1.0000 ng/ml

Temperature: 37.0 °C

Curve Offset:

None

Ascending

Descending

Delta Abs

Default


Save As OK Cancel Help

In this example, we will use the modified default parameters temporarily. To do so, we confirm the changes by clicking on the OK button. In this case, the selected parameters remain only valid until you leave the Kinetics/Time application or select a new method.

You may save the modified default method parameters permanently under an appropriate method name by clicking on the Save As button. In this case, the next time you use the Kinetics/Time application, just select your saved method from the Method menu or the method drop-down list box and you are ready-to-go using your individual parameters. You may change your method parameters at a later time or save the method under a different name.

We have now set up the Kinetics/Time application in a way to fit our needs by modifying the method, in this case the <Default Method>, and your window should now look like the one shown below.

Figure 25 Window with Kinetics/Time Application and All Changes Made



The following three things have changed when we closed the Method window. First, the Method Parameter List has been updated. Notice that it displays the parameters we have changed earlier in the method. Second, the dAbs/min column has been added to the data grid (for each sample, if we would have multiple samples).

We are now ready to start the kinetics run using the customized parameters. But before we do that, we need to blank the system on the selected wavelengths (the analytical and the background wavelength).

Blanking

A blank is required when lamps are turned on or certain parameters change (see Blanking Method for more information). In these cases, the instrument will remind the user with the message BLANK REQUIRED. Insert an empty cuvette or a cuvette with a blank solution in the cell holder at the measurement position and close the sample compartment.



Now click the BLANK button. The *Status & Control Frame* displays “Blanking ...” during the blanking and “Last blanked at 11:13 AM” when it is complete. In this example, two blank readings are taken - one at 540.0 nm and one at 480.0 nm.

The system is now ready to provide accurate readings.

Reading

Insert a cuvette containing a sample into the Single Cell Holder (or the appropriate position(s) of the Multi-Position Cell Holder) and close the sample compartment.

To simulate a kinetic reaction, we need a sample that changes absorbance over time. For this example, we will add a small amount of coffee creamer (any brand will do) to the top of a cuvette filled with water. Do not shake the cuvette. The dispersion of creamer will cause a change in absorbance over time. Begin the run as soon as possible after adding the creamer. Don't forget to close the door of the sample compartment before you start the run.

Also, make sure to click the *Dynamic Autoscaling* item in the *Axis* menu to ensure that you will see the kinetic data in the graph window during the run.




Now click on the READ button to start the kinetic run. As determined in the method, this will take approximately 60 seconds. During the run, the absorbances and the calculated rates for each time point (dAbs/min) are displayed in the data grid. You should also see the net absorbances in the graph window. After all readings have been taken within the run, the *Rates* item in the *Mode* menu will be enabled.

You now have three modes available and the windows should look like the ones below:


1. The *Data Acquisition mode*, which we have just completed.

Figure 26 Data Acquisition Mode



- The *Rates* mode, which shows the rate(s) or result(s).


Figure 27 Rates Mode



- The *Raw Data* mode. We can switch back and forth between the modes to display the acquired data or the calculated rate(s) or result(s).

After we have finished the kinetic run, we must decide what to do with the acquired data and results. We can discard them, print them, and/or save them. In any case, to complete the current run, select another application or method, or proceed with something else, we must execute *Save* and/or *Clear*.

Print, Save, and Clear



Print - To print the acquired data, select *Print* from the File menu in the Applications Frame.



Save - To save the acquired data, select *Save As* from the File menu in the Applications Frame or click on the *Save File* icon button.

The *Save File* dialog appears, which lets you enter an appropriate filename. Click on the OK button to save the data or Cancel to abort. If you don't change the target directory, the data file (.DUX) is saved to the following location "C:\DU800 Data\Generic User\Kinetics\



Clear - To end the current task and reset the application (e.g.; for another task or to leave the application), select *Clear* from the File menu in the Applications Frame or click on the *Clear* icon button. A message will inform you if the data has not been saved. The data grid and the graph are cleared, any unsaved data is discarded, and the application is reset.

You may then start another *Kinetics/Time* run or select another application.

System Software

General

The CD-ROM includes the *DU 800 System and Applications Software* setup files, the installation and operating instructions and all installation instructions for accessories as Adobe Acrobat (PDF) files. This section focuses on *System Software* and describes the user interface and generic functionality. The applications are described in *Applications Software*.

The software has been optimized for a screen resolution of 800 x 600 pixels or higher as well as the use of "Small Fonts". With a lower resolution, the main window will not fit on the screen and the selection of "Large Fonts" or "Other" will result in the misalignment of controls.

Default Directories

By default, the software will be installed in the *Program Files* path or the path specified during setup. The default installation directory is "DU800 System". Subdirectories and files that are created and kept beneath this directory, such as methods, etc., are considered system-related and not data-related.

Directory Level	Description
C:\	
Program Files	
DU800 System	INSTALL Root
Audit Trail	For Audit Trail files
DU800 Audit Trail.log	
System Backup Files	For Backup/Export files
My System backup.bkx	
Custom Applications	Holds custom applications
Diluted DNA.mth	
Methods	
Generic User	Methods Directory, Generic Level
Fixed Wavelength	
My Method #1.mth	
My Method #2.mth	
Wavelength Scan	
My Method #1.mth	
Administrator	Methods Directory, Administrator Level
User Name #1	Methods Directory, User Level
User Name #2	Methods Directory, User Level
Executable	Includes Program Files (don't touch)
Help	Includes HTML Help File(s)
PDF	Adobe Acrobat PDF Files

Data files that include readings, results, and scans are located in the data root or DATA directory, which is separate from the installation directory. As an exception, backup files for methods are located in the "Backup Files" directory of the DATA directory. The default directory for data files is *C:\DU800 Data*. The directory location may be changed by the administrator.

Directory Level	Description
C:\	
DU800 Data	DATA Root
Performance Validation PV Results 000818.txt	Holds Performance Validation result files
Generic User	Data Directory, Generic Level
Backup Files	For Backup/Export files
My Methods Backup.bkx	Holds backup methods
Fixed Wavelength	Applications-specific directory
My Data.dux	
My Data #2.dux	
Wavelength Scan	Applications-specific directory
My Data #1.dux	
Diluted DNA	Custom Applications Directory
My Data #1.dux	
Administrator	Data Directory, Administrator Level
Backup Files	For Backup/Export files
User #1	Data Directory, User Level
Backup Files	For Backup/Export files

File Extensions

The following represent valid file extensions:

File Extension	Description
MTH	Methods Database
DUX	Scans, Data and Result files
BKX	Backup/Export files
TXT	Performance Validation Result files
LOG	Audit Trail files

Passwords

The Administrator password provides access to the administrator functions. The default administrator password is "admin". This initial password should be changed by the administrator after installation.

The administrator sets up and maintains user accounts with the *User Account Manager* and establishes initial user passwords. From this time on forward, the user updates his/her own password without the involvement of the administrator. See *User Account Manager* and *Change Password* for more information.

The *Service Diagnostic* programs are password-protected and are only available to Beckman Coulter service personnel.

Logon Levels

The DU 800 includes three levels of operation. In all three levels, all available applications are fully accessible using the default settings.

Generic or <Generic User> - This is the default level and provides no methods/data protection and no user customization.

User - A user may log on with a user name and password. The initial password is supplied by the administrator. Only methods belonging to the individual user are shown and data are accessed and stored in user-specific directories. Each user can set and use his/her own preferences in User Options.

A group of users, such as personnel of a specific lab, could also be considered as a single user. Simply choose an appropriate name and share the password with the group.

Administrator - The administrator may log on as "Administrator" and the administrator password. Only methods belonging to the administrator are shown and data are accessed and stored using the administrator subdirectory. The administrator can set the system preferences in System Options. The system options apply to the <Generic User> and serve as the default preferences for the User Options.

See User Logon / Change User for more information.

On-Line Launch (with System Initialization)

The DU 800 Spectrophotometer is initialized by the Windows software when it is detected on the USB port. It is recommended to turn the instrument on before launching the software.

NOTICE


Lift the sample compartment cover of the DU 800 Spectrophotometer and verify that there is nothing in the sample compartment that can block the light beam.

After turning on the DU 800 Spectrophotometer for the first time, wait approximately 7 seconds before starting the software. Failing to do so, may result in the message "Another program is already running" or the system just sits there without initializing. In this case, close the software, turn the spectrophotometer OFF, wait a couple of seconds, and then turn it ON again. After waiting for approximately 7 seconds, start the software.

When an instrument is detected on the USB port, the software downloads the firmware to the DU 800 and builds the wavelength calibration table. This takes approximately 3 minutes. Then, various Startup Diagnostic Tests are performed and the result of each test is reported. The time for each of these tests varies. Please refer to Troubleshooting, Startup Diagnostic Tests should one or multiple tests fail.

During the initialization process, the window below displays the progress and allows the user to take appropriate action. The operator may click on the Abort button to end the system initialization.

Figure 28 DU 800 System Initialization Screen



The Print button, which can be used to create a startup diagnostic report, becomes enabled after the system has been initialized. To continue and operate the software, press Enter or click the Continue button after the Startup Diagnostic Tests have been completed.

If the software detects a spectrophotometer but cannot establish communication after the firmware has been downloaded, the user will be alerted by the message "Communication with the DU 800 UV/Vis Spectrophotometer could not be established". In this case, the software continues but without initializing the spectrophotometer (Off-Line Launch). Check the connection or see Section 7. Troubleshooting.

If the software detects a spectrophotometer on the USB port, the DU 800 System will be initialized automatically, if required.

After the system has been initialized and communication with the spectrophotometer is lost, for one reason or another, the user will be alerted by the message "Communication with the spectrophotometer has been lost. Please check the connection". The software re-establishes communication with the spectrophotometer automatically when the connection is restored.

Off-Line Launch

When no instrument is detected on the USB port, the software is launched in off-line mode.

This would be the case when the software has been installed on another PC for post-run analysis or other purposes. The fact that the software is running in off-line mode is indicated in the software caption.

Common Front End

User Interface


The DU 800 System and Applications Software integrates all system functionality and the various applications in a common user interface (front-end). This allows the user to operate all aspects of the system from a single vantage point. The integrated user interface includes key control elements, common to all applications.

The front-end and other system-related functions are described in the System Software section.

The standard applications are available with the *DU 800 System and Applications Software* installation. Standard and optional applications are described in the Applications Software section. Optional applications or software options can be added via a license contained on a floppy disk. The addition or installation of software options is described in Add Software Options.

The desired application is selected from the applications menu or a drop-down list box. When selecting or initializing an application the <Default Method> will be applied and the *Applications Area*, which includes the *Applications Menu Bar* and the *Data/Scan Area*, will reflect the functionality of the application as described in this section.

Figure 29 Integrated Frame-Oriented User Interface



Some commands and functions, such as Blank, Read, Stop, Save File, Open File, etc., apply to all applications and are described in Toolbar and Icons.

Main Menu - The top menu provides access to system-wide functions.

Method Parameters Frame - This list box displays relevant method parameters at all times.

Status & Control Frame - Displays the status of the system (User Logon, Blank Status, UV Lamp Status) and provides common control functions.

Toolbar and Icons - The toolbar at the top of the screen, just below the main menu, provides quick access to key elements related to the operation of the system and application.

Applications Frame - The applications frame holds the applications menu bar, all elements related to the selected application, and displays the acquired data.

Permanent Status Bar - The status bar at the bottom displays system parameters and messages.

Main Menu


The main menu provides access to all functions.

System Applications Methods Tools Diagnostic Help

During an "open run", functions in the *Applications* and *Methods* items are disabled. They will be enabled again when the user executes the *Clear* command, which "closes" the run. To switch to another application or select another method, the user must first close any open run. When the instrument is busy measuring samples, functions that may interfere are disabled temporarily.

Toolbar and Icons

The toolbar is fixed and cannot be removed by the user. It provides shortcuts for a number of common functions, such as BLANK, READ, and STOP buttons. In addition, other common functions can be accessed from the toolbar.



The top drop-down list box in the toolbar contains the *Applications* and provides the same functionality as the Applications menu item. The lower drop-down list box in the toolbar contains the *Methods* and provides the same functionality as some of the Methods menu items.

BLANK, (GO) READ/SCAN, and STOP Buttons



Blank, Read or Scan, and Stop are the most common commands and are executed through large icon buttons.

Read is used as the caption for the GO button for all applications, except *Wavelength Scan*. The *Wavelength Scan* application uses Scan as the caption for the GO button. Blank will blank on one or multiple single wavelengths (read blank) or a wavelength range (scan blank), depending on the application.

During data acquisition, the Stop button is enabled and halts the measurement when the current readings for the selected wavelengths have been completed or at the next possible stop point for a scan. Complete scan data that have been acquired will be retained and the user makes the choice to save or discard the data.

During an "open run" and "data acquisition", certain controls are disabled and cannot be used in order to protect the process.

The start of an "open run" is the measurement of the first sample. The run will be "closed" when the *Clear* command is given, usually after the data has been saved. The time in between is defined as an "open run".

For a definition of "open run", "data acquisition", and other terms, please refer to Terminology.

RediRead and RediScan



Opens the RediRead or RediScan window, which allows the user to take quick readings or scans. These icon buttons are shortcuts for the *RediRead* and *RediScan* items in the *Tools* menu.

Cell Match



Opens the Cell Matching window, which allows the user to match multiple cells. This icon button is a shortcut for the *Cell Matching* item in the *Tools* menu.

Edit Method

Opens the Method window for the current application with the currently selected method name and its parameters. This icon button is a shortcut for the Create/Edit Method item in the *Methods* menu.



Open File, Save File, and Clear



The Open File, Save File, and Clear buttons represent shortcuts for the File menu items of the current application and perform the same function as the respective items.

Opening a file is not like opening a file in Microsoft Word, there are important considerations in this software, which are described below.

To keep the structure clear and simple, the following has been implemented:

- a. Users cannot select another *Application* or *Method* and/or change the user during an "open run". The respective menu items and list boxes are disabled during an open run.
- b. The *Save File* and the *Clear* items/buttons are disabled when no data has been acquired or loaded. When data has been acquired or when a file has been loaded, the user must press *Clear* to end the "open run" or "file loaded" mode. *Save File* will save the current data but will not close the run. *Clear* will close the open run and reset the application. If necessary, it will remind the user to save data before clearing.
- c. Depending on the application, the *Open File* item/button may be disabled during an open run.

- d. All data files are associated with an application. When opening a file, the associated application is launched automatically, if required.

When a file has been opened, the following happens:

1. The methods drop-down list displays the filename instead of the method. This clearly indicates a "file loaded" mode instead of the regular "read mode". The displayed parameters are the file parameters, not the method parameters, and there is no link to a method name.
2. *Applications* and *Methods* menu items and list boxes will be disabled. The *Edit Method* icon button is enabled to allow post-run analysis. It also allows the user to create a new method based on the loaded file parameters.
3. Parameters stored within the file will be displayed in the *Method Parameters* list box. However, the "Method Parameter" column will be renamed to "File Parameter".
4. Measurement controls, such as *READ* or *SCAN*, are used to take readings or scans. These controls are disabled when a file has been loaded.
5. In the *Method* window, parameters such as Analytical Wavelength, etc. are locked when a file has been opened because they are related to measured data. Parameters that remain open can be modified to perform a post-run analysis. The data may then be saved under a different filename.
6. **NO MEASUREMENTS CAN BE TAKEN WHEN A FILE HAS BEEN LOADED.**

When opening a file or saving data, the initial directory is the applications-specific directory under the DATA root. Each logged-on user has his/her individual directories. These data directories are created automatically within each user directory and will be accessed by default for opening and saving data files.

The default directory for data is the user's directory in the DATA root.

If the DATA root is an UNC path and the system cannot connect to the network, the user will receive the message: "Cannot currently connect to the network! Do you want to use the local drive temporarily? [Yes] [No]. If yes, the default root or program directory will be used.

NOTICE

Existing DUX files cannot be overwritten. Trying to save a file under the same name will prompt the user to specify a new filename. When a file is saved by a user with Signature Privilege, the Add Electronic Signature dialog is displayed.

Accessories

Accessories

Opens the Accessories window and allows the user to setup the accessories for the system. This icon button is a shortcut for the *Accessories* item in the *System* menu.


Method Parameters Frame

Method Parameter	Value
Sampling Device	Holder
Holder Type	6-Cells
Number of Cells	6
Read Mode	Abs
Calculation Mode	None
# of Wavelengths	3
Wavelength 1	200.0 nm
Wavelength 2	250.0 nm
Wavelength 3	300.0 nm
Read Average Time	0.5 sec

This part of the front-end provides a listing of the major method parameters. With an open file, the file parameters are displayed and the column header changes to "File Parameter".

The data in the grid are for information only and cannot be edited.

Status & Control Frame



This part of the front-end combines important status information with the control of key spectrophotometer elements, such as sources and sampling device.

Status Information

User Status

The current user is displayed in the *User Status* label. When no user is logged on, the system is in the <Generic User> mode. The above frame shows the system in generic user mode.

Blank Status

The blank status is shown in the *Blank Status* label and provides the time for the last blank. Initially, it reads BLANK REQUIRED and when the blank flag has been reset by the program. Each time a parameter changes that can affect the reading, the blank flag will be reset.

During the time it takes to read a blank, the label reads Blanking ... and **Last Blanked at <time>** when a blank has been obtained. The time format depends on the international settings of Windows.

UV Lamp Status

When the UV lamp is off, the UV lamp status label is empty. During the warm-up phase of the UV lamp, the *UV Lamp Status* label reads Warming up UV lamp ... and the icon on the respective source button flashes slowly.

After the warm-up, the label reads **UV Lamp On for <time>**.

Status & Control Elements

Sources

The Visible and the UV sources are represented by icon buttons. They provide both, feedback and control. The sources can be turned on or off independently using the icon buttons or the items in the system menu. The visible lamp is executed instantly. The UV lamp requires a minimum warm-up time of 30 seconds; however, a 30 minute warm-up time is recommended.

Holders

The currently selected cell holder is displayed in the Status & Control Frame. Buttons are used for the available cell positions and HOME. The current cell position is indicated by a button in the pressed state. With a transport, the HOME button is shown. HOME re-aligns the transport and moves the transport to a fixed location. Notice that HOME is not a measurement position.

The user may control the transport or cell positions using the cell buttons when "Manual" has been selected as the *Number of Cells Used* in the method. Single cell holders are represented by a single position.

The Cell Holder elements are as follows:

Any Cell Holder on a Static Mount



An Auto 6-Cell Holder on a Transport



Any Single Cell Holder on a Transport



An Auto 12-Cell Holder on a Transport



Applications Frame

The applications area holds the applications menu and all elements related to the current application, such as the grid for data and results, graphed wavelength scan data, etc.

Figure 30 Applications Frame

The screenshot shows the Applications Frame with three absorption values displayed at the top: Abs @ 200.0 nm (0.9820), Abs @ 250.0 nm (0.8914), and Abs @ 300.0 nm (0.0762). Below these is a table with the following data:

Sample ID	Abs @ 200.0 nm	Abs @ 250.0 nm	Abs @ 300.0 nm
Sample 1-1	0.7255	0.5534	0.5995
Sample 1-2	0.3096	0.3219	0.7947
Sample 1-3	0.0340	0.7807	0.8345
Sample 1-4	0.7290	0.0654	0.4340
Sample 1-5	0.8826	0.8105	0.3935
Sample 1-6	0.9820	0.8914	0.0762
Sample 2-1			
Sample 2-2			
Sample 2-3			
Sample 2-4			
Sample 2-5			
Sample 2-6			

The appearance of the applications frame is dependent on the respective application. The applications are described in the Applications Software section.

Permanent Status Bar

The status bar is located at the bottom of the screen. It displays the status of the operation and the key spectrophotometer parameters. The status bar is fixed and cannot be removed.

Ready	Cell Match OFF	600.0 nm	0.308 Abs	24.8 C	10 Jul 2001 09:38 AM
-------	----------------	----------	-----------	--------	----------------------

Operation and Status Message

The left field in the status bar is the operation and status message, which indicates the current status and/or operation (e.g.; "Scanning cell #4 ...").

Spectrophotometer Status

The sunken labels represent spectrophotometer status information, which are: cell match, current wavelength of the drive, current Abs or %T, current temperature of the Temperature Controller, if any. With no Peltier Temperature Controller, "N/A" is displayed.


System Menu



User Logon / Change User

The <Generic User>, which is the default, is not considered a logged-on user. Therefore, in the <Generic User> mode, the menu item reads *User Logon*. When logged on as a user or administrator, the caption changes to *Change User* and the *Change Password* and the *Log Off* items become enabled.

Figure 31 User Logon Screen




The *Logon User* and *Change User* items perform identical functions, only the terminology is different.

Change Password

Each user is able to change his/her password at any time. *Title 21 Code of Federal Regulations (21 CFR Part 11) Electronic Records; Electronic Signatures* requires the periodic change of the user password. See System Options, Regulatory Tab for the necessary setup.

Figure 32 Change Password for Administrator Screen



The screenshot shows a dialog box titled "Change password for Administrator". It features three text input fields for password entry: "Enter old password:", "Enter new password:", and "Confirm new password:". At the bottom of the dialog are two buttons: "OK" and "Cancel".

The OK button is disabled until the number of characters in the *Confirm New Password* field is identical to the *Enter New Password* field and the new password is confirmed in the *Confirm new Password* field.

A user account must be established by the administrator before a user can log on. The User Account Manager, which allows the creation of user accounts, is strictly an administrator function.

Log Off

The *Log Off* item returns the system to the <Generic User> mode and the *Fixed Wavelength* application.

Sources

The menu items *UV Source* and *Visible Source* allow the operator to turn the lamps on and off. The respective icon buttons in the Status & Control Frame perform the identical functions. Check marks are used to indicate the status of the lamps.

Accessories

The *Accessories* item allows access to the setup window for the system accessories. The installed accessories must be defined for the system to function properly. After making any hardware changes to the DU 800, such as changing the cell holder or installing a sipper, the appropriate tab of the accessories window must be updated.

Accessories

Shortcut for Accessories.

For more information on the Accessories setup, refer to:

- Transport and Holders
- Sipper Accessory
- Batch Sampler

- Peltier Temperature Controller

Exit

Exits the *DU 800 System and Applications Software*. Before exiting, the system will prompt the user when data has not been saved.


Transport/Holder

Depending on the mode, the software automatically detects the presence and type of the Transport. The proper selection is made automatically in the Transport/Holder tab. The user must still select the *Holder* that is mounted on the transport.

On-Line Mode: Transport is detected automatically; user cannot select it.

Off-Line Mode: Transport is set to “None”; user cannot select it.

Figure 33 Accessories Screen with Transport/Holder Tab Selected



The current Cell Holder must be defined by the user. Based on the Transport, the compatible cell holder items are enabled and the installed cell holder can be selected. For example, a multi-position cell holder can only be selected if a transport is installed or has been selected. The *Tm Microcell Holder* can only be selected if the *High Performance Transport* is installed and selected.

See Sampling Accessories, Cell Holders for additional information.


With the selection of the Custom Cell Holder, the custom cell holder frame becomes visible and the position for each cell location can be defined in millimeters. A custom cell holder can have a maximum of 12 positions. The cell locations must be within the range of 0 to 112 mm.

The Manual Position Control determines the cell position in mm. The *Enter* key or the *Apply* button will execute the command and move the transport to the specified position. The input range is from 0 to 112 mm.

Sipper

The *Sipper Accessory* uses a peristaltic pump to aspirate the sample into a flow cell for measurement in the DU 800. The sipper is used to semi-automate the task of measuring a large number of samples. The Sipper Installed item must be checked when a sipper is installed.

Figure 34 Accessories Screen with Sipper Tab Selected



Fill Time - The fill time is the amount of time, in seconds, that the pump runs to aspirate a sample into the flow cell. The fill time also determines the amount of sample that is aspirated. The more viscous the sample, the longer the pump will need to fill to aspirate a given volume. The recommended minimum sample volume with a sipper is 1.0 mL.

If the sample will be returned after analysis, to minimize contamination, the sample solution should fill the inlet tubing, the flow cell, and the outlet tubing, but not enter the pump tubing.

If the sample volume is critical, the amount of time required to aspirate that amount of sample should be checked periodically.

Return Time - The return time is the amount of time, in seconds, that the pump returns the sample when [RETURN] is pressed. It is normally set to equal or slightly greater than the fill time.

Read Delay Time - The read delay time is the amount of time, in seconds, that the sample is in the flow cell after the pump stops and before the reading is taken. This allows the sample to stop flowing and to stabilize. A settling time of at least 1.0 second is recommended.

Speed - The speed of the pump is controlled with the speed parameter. A faster pump speed will aspirate samples faster. For typical aqueous samples, a pump speed of 8 is recommended. Viscous samples may require a slower pump speed, because of the surface friction developed when moving through the tubing. Organic samples with a low vapor pressure may require a slower speed to reduce evaporation. Reduce the pump speed if splattering occurs when samples are returned.


See Accessories, Sipper Accessory for additional information.

Batch Sampler

The *Batch Sampler* is used to fully automate the measurement of a large number of samples.

The Raise Arm, Advance, Lower Arm, Arm to Wash, and the Arm to Sample buttons may be used to control the batch sampler manually.

Figure 35 Accessories Screen with Batch Sampler Tab Selected



Batch Sampler - Select the available batch sampler or 'None'.

Pump Parameters - The parameters Fill Time, Return Time, Read Delay Time, and Speed are sipper parameters and are described in the Sipper Tab.

Wash Time - If a wash between samples is selected, input the time for the wash.

Wash Between Samples - If the batch sampler has a wash station, a wash between each sample can be selected by checking the box.

Return Mode - Select the desired Return Mode by clicking on the respective radio button. The percentage values are that of the sipper fill time.

The return modes have the following uses and limitations:

Waste Sample - Flushes the sample to waste for 5 seconds.

No Return

- Raise arm
- Flush 5 seconds

Forward Flush - Returns most of the sample with the aspirator arm lowered, raises the aspirator arm to finish returning the sample, then flushes to waste to remove the maximum amount of sample from the tubing.

80% Return

- Raise arm
- Wait 1 second
- Return 20%
- Wait 1 second
- Return 20%
- Wait 1 second
- Flush 100%

Forward Flush will not aerate the sample, but has the possibility of creating aerosols and does send the sample through the pump tubing.

Reverse Flush - Returns all of the sample with the aspirator arm lowered, then raises the aspirator arm and attempts to remove the maximum amount of sample from the tubing.

200% Return

- Raise arm
- Wait 1 second
- Return 20%
- Wait 1 second
- Return 20%
- Wait 1 second
- Return 100%

Reverse Flush will not aerate the sample, but has the possibility of creating aerosols. It does not send the sample through the pump tubing.

Lower Flush - Returns all of the sample with the aspirator arm lowered.

200% Return

- Raise arm

Lower Flush will clean the tubing as well as Reverse Flush, but does not create aerosols and does not send the sample through the pump tubing. It does aerate the sample.

In general, if return of the sample is desirable:

- Use Forward Flush if aerating (bubbling) the sample causes problems
- Use Lower Flush if creating aerosols causes problems.
- Use either Reverse Flush or Lower Flush if the sample will degrade the pump tubing.


See Accessories, Batch Sampler for additional information.

Peltier Temperature Controller

The *Peltier Temperature Controller* is used in conjunction with the *High Performance Transport* and a *Peltier Temperature-Controlled Cell Holder*. It cannot operate if these two accessories are not installed.

The temperature controller is activated by checking the Temperature Controller Enabled box.

Figure 36 Accessories Screen with Temp Controller Tab Selected



Manual Control

The temperature controller is enabled by entering the desired temperature and checking the *Temperature Controller Enabled* box. The input range is from 10 °C to 90 °C. The range that is actually suitable for a 6-Position Cell Holder is from 20 °C to 40 °C.

Clicking the OK button will activate the temperature controller. If the temperature controller does not activate, the following warning message will be displayed: "The Peltier Temperature Controller could not be activated. Please verify that the Temperature Controller, High Performance Transport, and the Cell Holder are connected properly and that the controller is powered on".


Automatic Control

With *Kinetics/Time*, *Enzyme Mechanism*, and *Experimental T_m Analysis* applications, the temperature is controlled directly from the respective application. Temperature control will be initiated within these applications if the *Temperature Controller Enabled* box is checked and if all the requirements for Peltier temperature control are


met. The temperature range that is actually controlled depends on the cell holder. The *T_m Microcell Holder* has a much wider temperature range than other cell holders and significantly different control dynamics.

Applications Menu

Standard and Custom Applications



Standard, Optional, and Custom Applications



Standard and Optional Applications

The Application menu provides the user access to the standard applications and any optional applications that are installed. There are two types of optional applications:

Extended Standard Applications (e.g.; Wavelength Scan II)

The standard application (*Wavelength Scan*) is replaced by the optional application (*Wavelength Scan II*). The optional application expands the functionality of the standard application.

Stand-Alone Applications

These application items are added to the menu.

Standard applications are available immediately after software installation. The administrator may add optional applications that have been purchased. See Add Software Option(s) for details.

When selecting an application, all methods belonging to the application will be loaded and can be selected. The <Default Method> is the initial method. The related menu items and the methods list box will be rebuilt.

The applications drop-down list box is a shortcut and performs the same function.

All available standard and optional applications are described in Applications Software.

Custom Applications


A custom application is a method from one of the existing applications with a link to the respective application. A custom application may be used by any user to safely run a "canned" application. The method parameters of a custom application cannot be modified.

By default, this menu item is disabled. It is enabled with one or more custom applications. Regular methods can be copied to the *Custom Applications* and become a new "application".

A method-based custom application is associated with an existing application and protected from common method editing. The limit for the number of custom applications is 30.

Custom applications are publicly available applications and can only be deleted by the administrator.

Methods Menu



<Default Method>

Indicates if the default method is currently selected or allows the user to select it. The method drop-down list box in the toolbar provides a shortcut for this function.

Parameters of the <Default Method> can be modified and used temporarily but are not saved. The <Default Method> provides the initial parameters when creating methods. To save selected parameters, the method has to be saved under a new name.

Select Method


This item is disabled when there are no methods. Only methods that belong to the current user (<Generic User>, logged-on user, or administrator) will be available. The available methods are listed and can be selected in the *Select Method* sub menu. The current method name is checked.

The methods drop-down list box in the toolbar provides a shortcut for selecting a method. The first item in the list is always <Default Method>, followed by the available methods.

Create/Edit Method



This item opens the *Method* window for the current application and shows the method name and its parameters. The icon button *Edit Method* in the toolbar is a shortcut for this function.




There are multiple tabs for the selection of method parameters. The tabs and the content of the tabs depend on the current application. The method name appears above the tabs in bold (1). When the 'Save As' button is clicked, the user can enter a new name or modify an existing name. Clicking on OK or pressing Enter (2) saves the method under the given name. The new method name (3) will be available immediately in the Methods menu bar and the drop-down list box.

The first (left) tab is open by default and contains the key parameters for the selected application. Tabs with additional parameters may be present, depending on the application. The Sampler, Headers, and Info tabs are common to all applications. Additional information can be found in the Help and in the Software Reference Manual.

Delete Current Method

This menu item is disabled with the <Default Method> selected. Before a method is actually deleted, the following message prompts the user to confirm the action.


Figure 37 Delete Current Method Screen



Copy Methods from Other Users

The list box shows all methods from other users. One or more methods can be selected and copied to the current user by clicking on the Copy From User button. The Copy From User button is enabled when one or more items are selected.

Figure 38 Copy Methods from Other User Screen




The <Default Method>'s are fixed methods and are not listed. When identical method names are encountered, a sequential number ("<method name>-01") is appended to the copied method name.

Copy to Custom Applications

A method can become a custom applications by copying it to *Custom Applications*. This menu item is disabled for a <Generic User> but enabled when a user or

administrator is logged on. Only methods from the current user or the administrator are shown and can be copied. The <Default Method> is a fixed method and is not listed.

Figure 39 Copy Methods to Custom Applications Screen



One or more methods may be selected and copied by clicking on the Copy To Custom button. The button is enabled when one or more items are selected.


When identical method names are encountered, a sequential number ("<method name>-01") will be appended to the copied method name.

A custom application has an association with the application and is protected from common method editing. Custom applications are publicly available applications and can only be deleted by the administrator.

Backup/Export Methods

This function allows a user to backup one or multiple methods from the current user to a file. It can also be used to 'physically move' one or multiple methods from one instrument to another. The <Default Method> is a fixed method and is not listed.

Figure 40 Backup/Export Methods Screen




One or more methods may be selected and backed up by clicking on the Backup/Export button. The button is enabled when one or more items are selected.

Backup files have the extension ".bkx". The initial backup directory is the Backup Files folder in the DATA directory (e.g.; C:\DU800 Data\Steiner\Backup Files). It will be created automatically, if it does not already exist.

Restore/Import Methods

This function restores all or selected methods from a backup file. The initial backup directory is the *Backup Files* folder in the DATA directory (e.g.; C:\DU800 Data\Steiner\Backup Files). However, the user may browse and select a backup file from another location (e.g.; Drive A:).

Figure 41 Restore/Import Methods Screen




All methods in the backup file are shown and can be restored or imported. One or more methods may be selected and restored by clicking on the Restore/Import button, which is enabled when one or more items are selected.


The window features the Overwrite Existing Methods function. When this function is unchecked and an identical and existing method name is encountered, a sequential number ("<method name>-01") will be appended to the restored method name.

Tools Menu


Generic User



User Logon



Administrator Logon



Both, RediRead and RediScan provide an easy and fast way to take simple readings or scans. These modes act like "interrupt" programs that keep the workflow of the current application intact.

When a multi-position cell holder is used, the Cell Matching function can be used to correct for differences in the cell positions.

To further process acquired data or obtain an independent, readable file, you may want to make it compatible with a commercial program using the Data Export function. The currently displayed data may be saved as a CSV (comma separated values) file that is compatible with commercial spreadsheet programs, such as Microsoft Excel.

Files may contain one or more Electronic Signature(s). The user may view the electronic signature(s) in a file that has been opened or add a signature to an existing DUX file, if he/she has the necessary privileges.

The User Options allow a logged-on user to set his/her preferences.

The Administration item is only enabled when the administrator is logged-on. The administrator has access to the following menu items:

- System Options
- User Account Manager
- Delete Custom Applications
- System Audit Trail
- Source Scheduler
- Performance Validation Scheduler
- Backup/Export System
- Restore/Import System
- Add Software Option(s)
- Create Backup License File


RediRead

Redi Read

RediRead provides an easy and fast way to take readings at a fixed wavelength and provide a printed output. In this mode, the system will not store data or perform calculations. *RediRead* acts like an "interrupt" program that keeps the workflow of the current application intact.

The *RediRead* window is used to take single wavelength readings on one or more samples. The cell position of a multicell holder cannot be changed from within the program.

Figure 42 RediRead Screen



The read average time is fixed at 0.5 seconds. The user simply selects the appropriate read mode (Abs or %T) and wavelength.


Before taking a measurement, *RediRead* needs to be blanked on air or a blank solution. This is done by clicking on the Blank menu item or button. When a blank has been performed, the indicator BLANK REQUIRED will change to BLANKED and the user is ready to take measurements.

NOTICE

Because this mode is an independent program/module, the blank is owned by this mode and does not affect the blank of the application.

Trying to read a sample without a blank causes the following message to pop up. 'Yes' will read the sample without a blank.

Figure 43 RediRead Dialog Screen



The current wavelength and result are displayed in large characters, in addition to the grid. This makes it easy to walk up with a sample, take a measurement, read the result from the screen, and walk away.

The Read menu item or button takes a reading based on the selected read mode and wavelength. Clear will clear the grid, Print will print the content of the grid, and Close will close the window.


RediScan

Redi Scan

RediScan provides an easy and fast way to take one or multiple wavelength scans and provide a printed output. In this mode, the system will not store data or perform calculations. *RediScan* acts like an "interrupt" program that keeps the workflow of the current application intact.

Up to six wavelength scans can be taken displayed in an overlay fashion. The cell position of a multicell holder cannot be changed from within the program. To avoid overlaying scans, simply click on the Clear menu item or button before taking the next scan.

Figure 44 RediScan Screen



The scan speed is fixed at 1200 nm/min, which represents readings at a 1.0 nm interval. The default scan mode is *Abs* and the default wavelength range is from 200.0 nm to 800.0 nm. Use the *Set Scan Limits* item in the Axis menu to modify the wavelength range over which the sample is scanned.

The system options set by the administrator act as default parameters for the y-axis, the grid and the legends.


Before taking a scan, *RediScan* needs to be blanked on air or a blank solution. This is done by clicking on the Blank menu item or button. When a blank has been performed, the indicator BLANK REQUIRED will change to BLANKED and the user is ready to take a scan.

NOTICE

Because this mode is an independent program/module, the blank is owned by this mode and does not affect the blank of the application.


Trying to take a scan without a blank causes the following message to pop up. 'Yes' will scan the sample without a blank.

Figure 45 Scan Without a Blank Dialog



The Scan menu item or button takes a wavelength scan based on the selected mode and wavelength range. Clear will clear the scan window, Print will print the current scan(s), and Close will close the window.

Function Menu



Trace

Trace is used to display the graphed data at the trace line position of the x-axis.

The menu item is disabled when there is no data. By default, the Trace function is inactive. When selected, the cursor in the scan window becomes a two-pointed horizontal arrow. As the trace line is moved across the screen using the mouse, the displayed values will be updated. In the case of curve fits, the actual data point values will be displayed when the trace line is near a data point.

Within the scan window, the left mouse button alternately removes and activates the trace line. In the active state, the cursor is limited to the scan window.

The data is shown in the respective color.

Trace is de-activated by selecting another function, such as *Default (Off)*, *Annotate* or *Zoom*.


Annotate

The Annotate menu item is disabled when there is not data. When the Annotate menu is activated or checked, the cursor in the graph window becomes an “A” character.

This function applied to graphic areas. By clicking on a location within the graph window, a realizable box opens in which text can be entered. The text will be fixed to the location when Enter is pressed.

By clicking on a location within the graph window, a realizable box opens in which text can be entered. The text will be fixed to the location when Escape is pressed.

The annotations and its locations are retained.



Within the annotation box, a right mouse click brings up a popup menu, which allows the user to fix, delete, or change the appearance of the annotation.

An existing annotation can be moved by holding down the *Ctrl* key and clicking on or near the text with the left mouse button. Releasing the mouse button will fix the annotation again.

Clicking on an existing annotation opens the annotation box and the annotation can be modified. The annotations and its locations are retained and saved with the scan. The number of annotations is not limited.

Annotate is de-activated by selecting another function, such as *Default (Off)*, *Trace* or *Zoom*.

Zoom

The feature can be used to expand any portion of the graph.

The menu item is disabled when there is no data. When selected, the cursor in the scan window becomes a magnifying glass. To zoom in, click the left mouse button and drag the mouse to another location. When releasing the mouse button, the system zooms in onto the rectangle that has been drawn. The zoom can be repeated to zoom in further.


A double-click on the right mouse button will zoom out.

Zoom is de-activated by selecting another function, such as *Default (Off)*, *Trace* or *Annotate*.

Grid

The default of this item is based on the system or user options setting. The grid can be turned on or off. This setting is temporary and will not be retained.

Axis Menu




Set Scan Limits

This item brings up a window, which allows the user to set both the x- and the y-axis parameters manually.

The Y-Axis limits are based on the current read mode. To modify the current wavelength range, change the X-Axis Limits.

Figure 46 Set Scan Limits Screen



When OK is clicked, the graph window will be refreshed.

Autoscale

The *Autoscale X*, *Autoscale Y*, and *Autoscale Both* items are enabled when data is available. These items automatically and immediately scale the respective axes.

Dynamic Autoscaling

This item is checked or unchecked depending on the system or user options setting. Dynamic autoscaling means that the y-axis will be re-scaled during data acquisition. The menu setting is temporary and will not be retained.

Read/Display Mode

The Abs and %T items indicate the current read mode by a check mark. When clicked, the scan(s) will be re-displayed immediately. The menu setting is temporary and will not be retained.

Cell Matching



When a multi-position cell holder is used, the *Cell Matching* function may be applied to correct for differences in the cells. Cell matching is performed using two or more cuvettes filled with a blank solution that are placed in a multi-position cell holder. The cell matching will be performed at the wavelength(s) selected. Up to three wavelengths can be used.

When the Match Cells button is pressed, the first cuvette is blanked and the value set to zero. Then, each following cuvette is measured and assigned a value relative to the first cuvette. These values are displayed in the grid as dAbs and d%T.

Figure 47 Cell Matching Screen

Cell	480.0 nm		520.0 nm		600.0 nm	
	dAbs	d%T	dAbs	d%T	dAbs	d%T
1						
2						
3						
4						
5						
6						

When the Use Cell Match box is checked, all readings taken by the instrument - in any application, except *Wavelength Scan* - will be corrected by the values displayed in the cell match window. This correction can be disabled and enabled without losing the correction values. The status bar indicates if the cell match correction is turned on or off.

NOTICE

When using cell match, make sure that the wavelength(s) are identical to the wavelength(s) of the application and that the cell match values are recent and from the same cells in the same position and that the cells have not been removed.

For the most accurate correction, the cell match readings should be taken at the same wavelength(s) as in the subsequent application and the cuvettes should not be removed from the cell holder before the analysis. If the cuvettes are removed, it is necessary to return each to the same cell position in the same orientation for the correction to be valid.

Cell match correction is recommended when using the *6-Position Microcell Holder* and the *12-Position Microcell Holder*.

The column labels indicate the wavelength and the matching values in dAbs and d%T. The number of cells displayed and processed here is based on the current cell holder selection in the Accessories window. The maximum number of cells is 12.

The selection of a different number of wavelengths or the change of a wavelength will clear the grid and a new match has to be performed.

Data Export

The *Data Export* item allows the user to save the currently displayed data as a CSV file. The CSV (Comma-Separated Values) format is a universal format and recognized by third party spreadsheet programs such as Microsoft Excel.

Figure 48 Example of Data Exported as a CSV File

	A	B	C	D	E	F	G
1	Data Acquired:			#####			
2	Company/Institute:			Beckman Coulter, Inc.			
3	Department:			Bioresarch Development Center			
4	Operator:			Steiner			
5	Comment:			This is a note			
6							
7	Product Name:			Abuprofin			
8	Lot #:			1234			
9	Read Mode:			Abs			
10	Calculation Mode:			Formula			
11	Number of Wavelengths:			2			
12	#	Wavelength					
13		1	250				
14		2	300				
15	Formula:			wl350 / wl400.0 x 1000			
16	Unit:			ng/ml			
17							
18		250.0 nm	300.0 nm	Results			
19	Sample ID	Abs	Abs	ng/ml			
20	My Sample ID #1	0.0374	0.0323	12.4939			
21	Sample 2-1	0.0374	0.0323	9.4832			
22							
23							

The CSV file is recognized by compatible programs and will load the data automatically for further analysis. To create a CSV file, simply click on the *Data Export* function in the *Tools* menu and the current data will be exported to a CSV file. The *Data Export* menu item is only available when data has been acquired or when a file has been loaded. Selecting this menu item will display a common *File Save* dialog.

The format of the CSV file follows these guidelines:

- Text information is embedding in quotes (to make sure intended commas are not interpreted as delimiters)
- All Header fields are text
- All Labels are text

Example File

```
"Data Acquired:",,,,"22 May 2000 03:50 PM"
"Company/Institute:",,,,"Beckman Coulter, Inc."
"Department:",,,,"Bioresearch Development Center"
"Operator:",,,,"Steiner"
"Comment:",,,,"This is a note"
"Product Name:",,,,"Abuprofin"
"Lot #:",,,,"1234"
"Read Mode:",,,,"Abs"
"Calculation Mode:",,,,"Formula"
"Number of Wavelengths:",,,,"2"
"#", "Wavelength"
"1", "250.0"
"2", "300.0"
"Formula:",,,,"wl350 / wl400.0 x 1000"
"Unit:",,,,"ng/ml"
, "250.0 nm", "300.0 nm", "Results"
"Sample ID", "Abs", "Abs", "ng/ml"
"My Sample ID #1", 0.0374, 0.0323, 12.4939
"Sample 2-1", 0.0374, 0.0323, 9.4832
```

Electronic Signatures

21 CFR Part 11

By definition, data and result files (.DUX) that are created by the DU 800 Spectrophotometer are electronic records. *Title 21 Code of Federal Regulations (21 CFR Part 11) Electronic Records; Electronic Signatures* came into effect on 20 August 1997 and sets forth criteria under which the agency considers electronic records, electronic signatures, and handwritten signatures executed to electronic records to be trustworthy, reliable, and generally equivalent to paper records and handwritten signatures executed on paper. People using electronic signatures must certify to the agency that the electronic signatures in their system are intended to be the legally binding equivalent of traditional handwritten signatures.

For more information, please refer to:

Code of Federal Regulations

Title 21, Part 11: Electronic Records; Electronic Signatures

U.S. Government Printing Office


<http://www.access.gpo.gov/nara/cfr/cfr-retrieve.html#page1>

(CITE: 21CFR11.1)

View Electronic Signature(s)

This item is enabled when a file has been opened and when an electronic signature exists in the file. The function calls a window that displays the Signer's Name and Role (Author, Reviewer, or Approver), the Date/Time of the signature, and a Note or Annotation (if one exists). All signatures included in the file will be displayed. A file may contain up to five signatures.

Figure 49 Electronic Signature Authentication Screen




Add Electronic Signature

The Electronic Signature consists of encrypted information based on the user name, and password as well as the serial number of the instrument. In addition to the signature, the file also includes:

- Name of the Signer (in readable form)
- Date and Time of the Signature
- Role of Signer (Author, Reviewer, or Approver)
- Note (empty by default)

Once added, the Electronic Signature will become a fixed part of the DUX file.

Figure 50 Add Electronic Signature Screen



The *Add Electronic Signature* menu item will give a logged-on user or the administrator the ability to sign a file. The menu item is only enabled when a file has been opened and if the user has a *Signature Privilege*, which is assigned by the Administrator in the *User Account Manager* screen. Before a signature can be added, the user will be prompted to repeat the logon. The signature is added to the existing file immediately. The file does not need to be saved to add the signature.

Existing DUX files cannot be overwritten. Trying to save a file under the same name will prompt the user to specify a new filename. If the user saves data under a new filename, all existing signatures will be removed.

A user or the administrator may sign a file as an Author, Reviewer, or Approver, depending on the available signature privilege. Only one Author is allowed per file. If an Author already exists, the Author button will be disabled. Multiple Reviewers and Approvers are permitted. When attempting to add a signature twice, e.g., same user in a different role, the user will be prompted. A file can include a maximum of five signatures.

The Note field allows the user to add an annotation.

User Options

This item is only visible when a user is logged on and disabled with a <Generic User> and the administrator. The function calls a window and lets the user set his/her preferences. Many user parameters are identical to the system parameters.

The default settings for user parameters are the system parameters. User preferences have priority over system parameters and are used in place of the system parameters. The parameters that are applied to the <Generic User> are the system parameters.

Output Tab

The Output tab includes the preferences for the screen and printer outputs. The content is identical to System Options, Output Tab.

Header Tab

The Header tab includes the custom header information and the selection for the length of the application header. This tab includes the *Department* and the *Operator* fields. Although, the content is different from the System Options tab of the same name, which has the system-wide *Organization* field, both tabs serve the same purpose (see System Options, Header Tab).

Graph Tab

The Graph tab includes the preferences for the graph window. The content is identical to System Options, Graph Tab.

Colors Tab

The Colors tab includes the color preferences for the graph window. The content is identical to System Options, Colors Tab.

Administration

The *Administration* item is only enabled when an administrator is logged on. Other users have no access to the functions included in this sub menu.

The following administrator functions are accessible through the *Administration* menu:

- System Options
- User Account Manager
- Delete Custom Applications
- System Audit Trail
- Source Scheduler
- Performance Validation Scheduler
- Backup/Export System
- Restore/Import System
- Add Software Option(s)
- Create Backup License File
- Security/Enhancements

NOTICE

Be aware that the Administrator and the User logon of the DU 800 are entirely independent of the Microsoft Windows Operating System logon.


System Options

In general, *System Options* apply to the <Generic User> and the Administrator. Many of these parameters serve also as default preferences for Users.

Setup Tab

The *Setup* tab determines the path of the user directory. In addition, it allows the administrator to determine timeout functions. This tab does not exist in the User Options.

Figure 51 System Options Screen with Setup Tab Selected



The information in the Administrator field will be used in message boxes that refer to the administrator. We recommend to enter the administrator's name. With an empty field, the text "Administrator" will be used.

Data files will be stored and accessed in the DATA Root, which is easily accessible through the Windows Explorer. The default directory is "C:\DU800 Data" and may be changed by the administrator. Instead of a fixed path, a UNC (Universal Naming Convention) path may be entered or selected.

When defining a new DATA root directory (complete path), the directory will be created automatically, if it does not already exist. Both, a local directory and an UNC

(Universal Naming Convention) path will be accepted. The Default button resets the parameters of this tab.

NOTICE

When defining a DATA root on a network drive, make sure that the required access privileges are available.


Instead of entering the path for the DATA root directory, you may browse for an existing directory by clicking on the Browse button to the right of the text field.

By default, the *Idle Timeout Functions* are unchecked.

The *Automatic Logoff* is required for 21 CFR Part 11 and will automatically log off the current user and return to the <Generic User> level when the system is idle for more than the specified time. When a user is logged on and the system is in an “open run” (e.g., it is busy with a Kinetics run), instead of logging off, the system continues the Kinetics run and displays a protected DU 800 screen saver.

The DU 800 screen saver is automatically selected whenever the DU 800 software is running and *Automatic Logoff* is selected. When the software is exited, the system will return to the previously selected screen saver. The DU 800 screen saver will lock the computer until the current user or the administrator enters his/her password.

Figure 52 User Authentication Screen



The *Elapsed Time* list box contains pre-defined entries: 1, 2, 5, 10, 20, 30 minutes and 1 or 2 hours. The default is 10.0 minutes.

If checked, the *UV Lamp Shutoff* and the *Visible Lamp Shutoff* items turn the respective lamps off when the system is idle for more that the specified time (idle means that no measurement has been taken for a while).


The respective lamp will not be turned off when the system is in an “open run” or a file has been opened. Also, the Source Scheduler has priority over the *UV Lamp Shutoff* and the *Visible Lamp Shutoff* functions.

Specify the *Elapsed Time* that should apply to the *Automatic Logoff* and the *Lamp Shutoff* functions.

Output Tab

The *Output* tab includes the preferences for the screen and printer outputs.

Figure 53 User Options Screen with Output Tab Selected



The default settings for the screen are MS Sans Serif, 8 pts and for the printer Courier New, 10 pts. The *Change Font* buttons call the Windows Dialog function for the selection of fonts.

NOTICE

A very small or large font size may not create the desirable output. In this case, you may consider going back to the default settings. The printouts are optimized for the default settings.

It is important to notice that the Screen Font Characteristics are only effective for the SCAN GRAPH and the DATA GRID areas - no other parts of the software are affected.

The # of Decimal Points determine the reported precision of the Abs and %T readings. The defaults are four (4) for Abs and one (1) for %T.


The *Highlight Results* option is unchecked by default. Initially, the Color Preference is set to "pink" and Bold is unchecked. The color is selected by calling the standard Windows Dialog function for the selection of colors. This option only applies to the RESULT column(s) in the DATA GRID areas and the printout.

The *Default* button resets the parameters of this tab.

Headers Tab

The *Header* tab includes the custom header information and the selection for the length of the application header. Headers apply to printouts and files.

Figure 54 User Options Screen with Headers Tab Selected



The Organization and Department fields are text fields. The information entered in the Organization field is applied system-wide. The department information serves as the default for the User Options and will be applied to the <Generic User> printout header.


Additional header information can be defined for each application (see Methods, Report Tab).

The Application Header allows the administrator or user to set the length of the application header for the printout. There are three priority levels, which are predefined.

Graph Tab

The *Graph* tab includes the preferences for the graph window.

Figure 55 User Options Screen with Graph Tab Selected



The Grid option turns the grid on or off. By default, this option is checked.

The Symbols option turns the symbols for data points on or off. By default, this option is checked.

The Emphasize Selected Scan option emphasizes the scan in focus with a bold line. By default, this option is checked.

The Dynamic Autoscaling option determines if dynamic autoscaling during data acquisition is performed or not. By default, this option is checked.

The Y-Axis Default values may be modified. The default values are: 0.0 to 3.0 for Abs (input range is -0.3 to 4.5) and 0 to 100 for %T (input range is 0 to 200).


The 3-D Offset frame includes the selection for the data shift to accomplish a 3-dimensional look for overlaid scans. The axis values apply to the first scan only. All other scans are shifted and the axis values cannot be applied. The default is 'Medium'.

The *Default* button resets the parameters of this tab.

Colors Tab

The *Colors* tab includes the color preferences for the graph window.

Figure 56 User Options Screen with Colors Tab Selected



This tab specifies the *Graph Colors* for both the screen and the printer. The colors are selected by clicking on the colored box. This evokes the Windows Dialog function for the selection of colors.

The items in the *Graph Colors* frame include the settings for the Window Background, the Graph Background, and the Grid.


The items in the Scan Colors frame determine the color of the individual scans.

The *Default* button resets the preferences of this tab.

Regulatory Tab

The parameters in the *Regulatory* tab also apply to *Title 21 Code of Federal Regulations (21 CFR Part 11) Electronic Records; Electronic Signatures*.

Figure 57 System Options Screen with Regulatory Tab Selected




The *Electronic Signature* feature is unchecked by default. When checked, it allows authorized users to add an electronic signature to a data or results file. The signature itself is based on the logon as well as other information and is unique (see section *Electronic Signature* for more information).

In order to restrict the use of the system to logged-on users and prevent unauthorized access, the *No <Generic User> Access* feature must be checked.

With the *Password Renewal* feature checked, the user and the administration passwords need to be renewed after a given time period. By default, this feature is disabled. When the feature has been enabled and the renewal time exceeded, the user must change his/her password in order to complete the logon process. The default renewal time is 90 days. The administrator may change this parameter when the feature is enabled. The entry is limited to 10-365 days.

The following message appears at logon, if the password requires renewal.

Figure 58 Password Renewal Screen



When continuing, the Password Renewal window lets the user update the current password. 'No' in the above message box or 'Cancel' from the Password Renewal window will abort the logon.

With the System Audit Trail feature checked, key events will be logged to an audit trail file. By default, this feature is disabled. This feature records all major events as indicated by the check boxes.

The following actions will be recorded, if checked:

- a. User Logon, including failed attempts, and User Logoff.
- b. Password change.
- c. Start and end of data acquisition (open run - start to clear).
- d. Loading and saving of a file.
- e. Saving methods (created or modified).
- f. Deleting methods.
- g. Copying methods (from another user or to custom applications).
- h. Backing up and restoring methods.
- i. Adding and deleting user accounts


The file has the fixed name "DU 800 Audit Trail.log" and is a protected file. The "master" System Audit Trail file is kept in the "Audit Trail" directory, located in the INSTALL directory. A "slave" System Audit Trail is maintained in the DATA directory to facilitate data archiving. This "slave" file is automatically updated each time a DU 800 data file is saved.

To keep the file at a manageable size, the operator will be prompted to inform the Administrator that the System Audit Trail requires maintenance. The message appears when the entered backup time of file size has been exceeded, whichever comes first. The range for the backup time is 0 to 2000 days and the range for the file size is 64 to 2048 Kbytes.

User Account Manager

The *User Account Manager* allows the administrator to add or delete user accounts and provide the initial password to an account.

Figure 59 User Account Manager Screen



The administrator is able to create (Add) and Delete user accounts. A user requires a password, which is initially given by the administrator. From this point on, the user is able to change his or her password at any time. The *Change Password* button allows the administrator to change the password of an existing user.

For each user, one or more Electronic Signature Privileges can be assigned. The default is 'Author'. The privileges include: 'Author', 'Reviewer', and 'Approver'. For more information on Electronic Signatures, refer to Electronic Signature(s).

The minimum number of characters for both user name and password is 4. The maximum length is 32 for the user name and 12 for the password. The current users are shown in a list box. To change a user's password or to delete the account, select the respective user name first.

To create a user account, simply enter the new user name and click on the *Add* button. The *Create Password* window pops up, which lets you enter the initial password. Enter a password and click on OK to return to the User Account Manager. You may then continue to add user accounts or close the window.


To delete a user account, select the appropriate user name in the list box and click on the *Delete* button. The user name will be removed from the list and the account deleted. The methods and data files for this user account will not be deleted.

Delete Custom Applications

The *Delete Custom Applications* menu item allows the administrator to remove custom applications. The menu item is disabled when a custom application is selected. To perform this function, a regular application must be selected.

The list box displays the current custom applications and lets the administrator select one or multiple methods with the usual Windows key combinations.

Figure 60 Delete Custom Applications Screen



Initially, the *Delete* button is disabled. It will be enabled when one or more items are selected. The deleted custom applications are removed permanently.

View System Audit Trail

The *System Audit Trail* menu item opens a new window and displays the content of the active audit trail file or a selected audit trail file. The file may be printed or backed up.

Figure 61 System Audit Trail Screen

Date	User	Event
March 27, 2002 9:49 AM	Administrator	Load file 'C:\DU800 Data\Administrator\Wavelength Scan\test1.dux'
March 27, 2002 9:49 AM	Administrator	Start of data acquisition
March 27, 2002 9:49 AM	Administrator	Data acquisition aborted
March 27, 2002 9:50 AM	Administrator	Add user 'mwong'
March 27, 2002 9:50 AM	mwong	Logon
March 27, 2002 9:50 AM	mwong	Save method 'Test Method FVWL'
March 27, 2002 9:50 AM	mwong	Save file 'C:\DU800 Data\mwong\Fixed\Wavelength\qwerty.dux'
March 27, 2002 9:50 AM	mwong	End of data acquisition
March 27, 2002 9:51 AM	mwong	Start of data acquisition
March 27, 2002 9:51 AM	mwong	Data acquisition aborted
March 27, 2002 9:51 AM	mwong	Data acquisition aborted
March 27, 2002 9:51 AM	mwong	End of data acquisition
March 27, 2002 9:51 AM	mwong	Logoff
March 27, 2002 9:51 AM	Administrator	Logon
March 27, 2002 9:52 AM	Administrator	Start of data acquisition
March 27, 2002 9:52 AM	Administrator	Save file 'C:\DU800 Data\Administrator\Kinetics Time\asdfghj.dux'
March 27, 2002 9:52 AM	Administrator	End of data acquisition
March 27, 2002 9:52 AM	Administrator	Load file 'C:\DU800 Data\mwong\Fixed\Wavelength\qwerty.dux'
March 27, 2002 9:52 AM	Administrator	Save method 'KT method'
March 27, 2002 9:53 AM	Administrator	Password changed for user 'mwong'
March 27, 2002 9:53 AM	mwong	Wrong password
March 27, 2002 9:53 AM	mwong	Logon failed

Buttons: Backup, Open, Print, Close, Help

When the *Backup* button is selected, the entire active audit trail is automatically saved in the "Audit Trail" directory, located in the INSTALL directory. A copy of this backup file is also created in the DATA directory.

With the *System Audit Trail* successfully backed up, the software will begin recording a new audit trail. The entries of the last two weeks will be retrieved to provide overlap and continuity.

Source Scheduler

The *Source Scheduler* allows the administrator to turn both sources on and off individually at a scheduled time for specified days.

Figure 62 Source Scheduler Screen

	UV			Visible		
	On	Off	Use	On	Off	Use
Sunday	OFF	OFF	<input type="checkbox"/>	OFF	OFF	<input type="checkbox"/>
Monday	07:00:00 AM	06:00:00 PM	<input type="checkbox"/>	07:00:00 AM	06:00:00 PM	<input type="checkbox"/>
Tuesday	07:00:00 AM	06:00:00 PM	<input type="checkbox"/>	07:00:00 AM	06:00:00 PM	<input type="checkbox"/>
Wednesday	07:00:00 AM	06:00:00 PM	<input type="checkbox"/>	07:00:00 AM	06:00:00 PM	<input type="checkbox"/>
Thursday	07:00:00 AM	06:00:00 PM	<input type="checkbox"/>	07:00:00 AM	06:00:00 PM	<input type="checkbox"/>
Friday	07:00:00 AM	06:00:00 PM	<input type="checkbox"/>	07:00:00 AM	06:00:00 PM	<input type="checkbox"/>
Saturday	OFF	OFF	<input type="checkbox"/>	OFF	OFF	<input type="checkbox"/>

Buttons: Default, Help, OK, Cancel

The default settings are shown in the above screen.

The Source Scheduler will not work when the system is in an “open run” or when a file has been opened.

The On and Off times can be set to OFF or from 5:00 AM to 10:00 PM for each day. The times are in ½ hour intervals and the format is based on the international setting of Windows.

The *Use* check box provides a quick way to turn the scheduler on or off for a particular day without changing the previously selected times. By default, these boxes are unchecked.

The *Default* button resets the preferences of the Source Scheduler.

Performance Validation Scheduler

The *Performance Validation Scheduler* allows the administrator to automatically execute one or more Performance Validation programs at a given day and time.

Figure 63 Performance Validation Scheduler Screen

	Start Time	Program or Group
Sunday	OFF	None
Monday	OFF	None
Tuesday	OFF	None
Wednesday	OFF	None
Thursday	08:00:00 PM	Run Group [1] Tests (3 min 30 sec)
Friday	OFF	None
Saturday	OFF	None

Results

Print Save

Results File Prefix: PV Test

Default Help OK Cancel

The *Start Time* can be turned OFF or selected in the drop-down list box for each day. The times are from 5:00 AM to 10:00 PM in ½ hour intervals. The time format is based on the international setting of Windows.

The Performance Validation Scheduler will not execute the selected test(s) when the system is in an “open run” or a file has been opened.

The Program or Group selections for the Performance Validation programs are:

- None
- Run Group [1] Tests (3 min 30 sec)
- Run All Tests (1 hour 4 min)
- [1] Wavelength Drive Test (24 sec)
- [1] Resolution Test (12 sec)
- [1] Baseline Flatness Test (1 min 40 sec)
- [1] Noise Test (1 min)
- [2] Stability Test (60 min)

BOTH THE VISIBLE AND THE UV LAMPS ARE REQUIRED FOR THE PERFORMANCE VALIDATION SCHEDULER.

THE SYSTEM WILL MAKE SURE THAT BOTH LAMPS ARE TURNED ON FOR THE TEST(S).


In order to receive accurate results, it is recommended that both lamps are warmed up (UV lamp requires one hour) prior to running Performance Validation tests. If either lamp is off at the scheduled time, the Performance Validation Scheduler will turn on both lamps and wait one hour before executing the selected test(s).

The Source Scheduler can be used to automatically turn on the lamps one hour before the Start Time selected in the Performance Validation Scheduler and to turn off the lamps at appropriate time (e.g., at the end of a work day).

After the selected Performance Validation program(s) have been completed, the results may be printed and saved automatically. By default, *Print* and *Save* are checked. If *Save* is checked, the results will be saved to a file. The filename is based on the Results File Prefix. The date of the test(s) will be appended to the filename prefix automatically in the following format: "<file prefix> YYMMDDHHMMSS.txt". The location for the file is the "Performance Validation" directory in the DATA root directory.


When the Performance Validation Scheduler activates, the following message is displayed:

Figure 64 Performance Validation Schedules Activation Dialog




When all tests are within specifications, the following message appears:

Figure 65 Performance Validation Scheduler Screen



With one or more specifications outside of the given limits, the following message appears:

Figure 66 Performance Validation Scheduler Screen with Error Message



The Performance Validation programs are described in Applications Software, Performance Validation.

The *Default* button resets the preferences of the Performance Validation Scheduler.

Backup/Export System

The function will be used by the administrator to *Backup/Export System* parameters. The reason for this is the safekeeping of system information in a file to restore the entire system after an system failure or when the *DU 800 System and Applications Software* must be reinstalled. The service representative may also use this function to backup the entire system and restore it after the service call.

First, the user selects a backup file in a Save Backup File common dialog box. The initial directory is "System Backup Files" in the INSTALL root directory and the initial filename is "Untitled.bkx".

The Backup/Export System function backs up all segments:

- a. User Accounts and Methods
- b. System and User Options
- c. Custom Applications


The installed applications are not included in the backup.

Restore/Import System

The *Restore/Import System* function restores selected segments from a system backup file.

First, the administrator selects a backup file in an Open Backup File dialog box. The initial directory is "System Backup Files" in the INSTALL root directory.

Figure 67 Restore/Import System Screen



By default, the *Overwrite* options are unchecked.

The following rules apply for restoring information:

1. Users Accounts and Methods

Non-existing user accounts: Users will be added with all passwords, methods, and user options retained.

Existing user accounts: With the *Overwrite* option unchecked, only non-existing methods will be restored. With the *Overwrite* option checked, all existing passwords, methods and user options will be replaced.

2. System Options

The system options will be restored and overwritten.


3. Custom Applications

With the *Overwrite* option unchecked, only non-existing custom applications will be restored. With the *Overwrite* option checked, all existing custom applications will be replaced.

Add Software Option(s)


The *Add Software Option(s)* function allows the administrator to install one or more software options.

Figure 68 Add Software Options(s) Screen




The license is initially provided on a floppy disk and is usually installed from there.

Figure 69 Select License File Screen



A dialog box will show the available license files on the selected medium. No other file types will be shown. The application(s) that are included in the selected license file will be displayed.

Figure 70 Add Software Options(s) Screen



Click on 'Yes' to install the software option(s). Installed software options cannot be removed. If the system encounters a file without a valid license, it will inform the user via a message box that this is an invalid license.

When the *DU 800 System and Applications Software* has been installed on a computer, which is not connected to a DU 800 Spectrophotometer, software options can be installed from the floppy disk that includes the license. However, it is important to notice that the software option must first be added to a computer with the DU 800 Spectrophotometer connected.

ATTENTION

Licenses for software options are provided on a floppy disk.

It is recommended to Create a Backup License File after one or more software options have been added. Make sure to store the file in a safe location.

Create Backup License File

The *Create Backup License File* function creates a backup file that includes all licenses for the currently installed software options. This function is recommended and should be performed after all software options have been added.

ATTENTION

Create a Backup License File after the software option(s) have been added.

MAKE SURE TO STORE THE BACKUP FILE IN A SAFE LOCATION.


In the case of an emergency, or if the software is re-installed on another computer, the backup license file allows you to restore the licenses for the purchased software options.

Security/Enhancements

Version 2.0, Build 74 includes new features for data security, 21 CFR Part 11 compliance, a customizable data directory structure, and another common enhancement. Only the administrator can activate these features, which will affect how logged-on users interact with the system. Some of these features are designed to give a company tighter control in regard to system usage, data acquisition and storage.

The new features of Build 74 are accessible through the "Security/Enhancement" item in the "Administration" menu.

Figure 71 Administration Menu.



Lock RediRead and RediScan


This feature gives the administrator the choice to lock RediRead and RediScan.

The check box for this feature is located in the "21 CFR Part 11 Compliance" tab of the "Security/Enhancements" item in the "Administration" menu.

By default, the parameter "Lock RediRead and RediScan" is unchecked. When checked, the menu items "RediRead" and "RediScan" in the Tools menu as well as the appropriate shortcut icons will be disabled for all users, including the administrators.

With the feature activated, RediRead and RediScan become inaccessible and cannot be used to take measurements.

Figure 72 21 CFR Part 11 Compliance.



Force Users to Save Data


This feature allows the administrator to force users to save the acquired data. With this feature activated, users cannot take measurements and then discard them. The acquired data must be saved to the specified location before exiting the application.

The check box for this feature is located in the "21 CFR Part 11 Compliance" tab of the "Security/Enhancements" item in the "Administration" menu.

By default, the parameter "Force Saving of Data" is unchecked. When checked: a) the Save dialog box cannot be cancelled and

b) the directory cannot be chosen (the default directories are used and the destination tree is displayed in the dialog box).


Figure 73 Sava Data.



Data/Methods Security

The window with the "Data/Methods Security" tab can be accessed from the "Administration" menu.

Figure 74 Data/Methods Security.



Disallow Creation of Methods by User

This feature allows the administration to set up the system so users cannot create their own methods and use them to acquire data and create records. In this case, only administrators will be able to create methods. Thus, limited users to methods that have been "verified" by the company.

This setting will have no affect on administrators.

The check box for this feature is located in the "Data/Methods Security" tab of the "Security/Enhancements" item in the "Administration" menu.

By default, the parameter "Disallow Creation of Methods by User" is unchecked. When checked, the

menu item "Create/Edit Method" and the shortcut icon "Edit Method" will be disabled for all users.

In addition, the following menu items will not be accessible with the feature activated:

- Create/Edit Method
- Delete Current Method
- Copy Method from Other User
- Copy Method to Custom Applications
- Backup/Export Methods
- Restore/Import Methods

Show Only Custom Applications

This feature works best in combination with the feature "Disallow Creation of Methods by User". With both features selected, the administrator can set up the system in a way that a user can only select and apply "Custom Applications". Thus, giving the company complete control over how data are acquired by limiting the system to Custom Applications, which are actually customized methods from the standard (and purchased optional) applications. In this case, the regular applications provided by Beckman Coulter will not show and can therefore not be accessed.

This setting will have no affect on administrators.

The check box for this feature is located in the "Data/Methods Security" tab of the "Security/Enhancements" item in the "Administration" menu.

By default, the parameter "Show Only Custom Applications" is unchecked. When checked, the standard (and purchased optional) applications that have been provided by Beckman Coulter will be hidden and only "Custom Applications" can be accessed by a user.

ATTENTION


When the feature is activated and the system cannot find a custom application, *Fixed Wavelength* is selected automatically as the application and the message "No custom applications. Please create one." appears. In this case, the administrator must first create a custom method and copy it to "Custom Applications".

Please shut down and re-start the *DU 800 System & Applications Software* for the feature to take effect.

Data Directory Structure

The window with the "Data Directory Structure" tab can be accessed from the "Administration" menu.

Figure 75 Data Directory Structure.



The "Default" directory structure is fixed as:

[Data Root Directory], as specified by the administrator

[User] - logged-on user

[Application] - selected application name

By default, all acquired data will be saved to and accessed from this location. e.g.:

C:\DU800 Data Generic User Fixed Wavelength

Alternatively, the administrator may set up the system to be using a customized directory structure by clicking on "Use Custom Settings". In this case, the fixed directory structure (or tree) will be replaced by a more flexible and customizable directory tree.

This, more flexible, data directory tree is defined as:

[Data Root Directory], as specified by the administrator

[Application] - selected application name

[Product Name*], or


[Component Name*]. or

[Lot Number*], or whatever the meaning of the appropriate entry is

[User] - logged-on user (optional)

* These parameters are part of the method setup ("Report" tab).

Figure 76 Data Directory Structure.



Level 1 of the subdirectory tree is fixed and represents the application, which is the application name. The application can be the name of a fixed application or the name of a "Custom Application".


Level 2 of the subdirectory tree can be any one of the following:

- Product Name
- Component Name
- Lot Number

These are method parameters from the "Report" tab that can be freely determined by the user - or the administrator who may provide the methods for the "controlled" system. When the selected entry is empty, the name Unknown Method is applied, giving a clear indication that a valid entry must be supplied.

With "Use Custom Settings", the parameter that has been selected as Level 2 will be disabled in "Headers for Current Run" so it cannot be overwritten by the user.

Figure 77 Headers for current run.



Headers for Current Run

Generic Header:

Department: Quality Control

Operator: Steiner, Ron

Product Name: Cyclobenzaprin

Component Name: Carbonate

Lot Number: A123456

Comment:


OK Cancel Help

Level 3 of the subdirectory tree is optional and is not used by default. When the parameter `Add [User] Level` is checked, an additional subdirectory with the name of the logged-on user will be applied.

Enhancements

The window with the "Enhancements" tab can be accessed from the "Administration" menu.

Figure 78 Enhancements



Security/Enhancements

Data Directory Structure | Data/Methods Security

21 CFR Part 11 Compliance | Enhancements

Automatically Convert to CSV File

Automatically Convert to CSV File

This feature allows the administration to set up the system so users CSV (Comma Separated Values) files are created automatically, without manually executing the "Data Export" function each time.

This setting affects all users and administrators.


By default, the parameter "Automatically Convert to CSV file" is unchecked. When checked, the "Data Export" item in the "Tools" menu will be disabled since the function is no longer needed.

With the feature activated, the CSV file will be created automatically when the 21 CFR Part 11

compliant DUX file is created. The CSV file will have the same filename as the DUX file, which is

specified by the user when the data is saved

Diagnostic Menu



The *Diagnostic* functions consist of two parts. The *System Status* and the *User Diagnostics* windows are accessible by any user and are described below. The *Service Diagnostics* functions are reserved for service personnel and can only be accessed through a specific password.


System Status

This window displays common information about the spectrophotometer and the operating system.

When the lamp time approaches 100% of the typical life expectancy, the administrator might consider replacing the respective lamp in order to avoid interruption when the lamp actually burns out.

The *Print* button outputs the information displayed in this window. The *Computer Info* button calls a separate Microsoft program which provides detailed information about the computer.

Figure 79 System Status Screen




User Diagnostics

Calibrate Wavelength

This function is used to calibrate the wavelength drive, if the wavelength accuracy is not within the expected range. The wavelength accuracy can be tested with the *Performance Validation* program.

Figure 80 Wavelength Calibration Screen




Performing a wavelength calibration takes approximately 2 minutes. The message "Wavelength calibration in progress ..." flashes during that time. When it is finished, the message reads "Wavelength calibration finished".

Run Baseline Test

The functionality of the *Baseline Flatness Test* takes a blank reading on air, followed by a sample scan on air to test the flatness of the baseline from 200 to 900 nm.

Figure 81 Baseline Test Screen



To test the baseline flatness, follow these steps:


1. Verify that both sources are turned on and that they have stabilized for at least one hour.
2. Verify that nothing is in the sample compartment that can potentially block or disturb the light beam.
3. Click on the Scan Baseline button or the respective menu item. The instrument automatically performs the test, displays the baseline and calculates the RMS Flatness. The expected RMS Flatness should be less than 0.001A.
4. You may click on the Print button or the respective menu item to print the baseline.

Set Scan Gains

The scan gains are set at the factory prior to shipment. The gains are checked and adjusted each time that the wavelength drive is calibrated, as described earlier. When a wavelength scan is performed using a blank that has significant absorbance in the

scanning range, it may be desirable to change the gain settings in the instrument to compensate for the absorbance of the blank. This will increase the dynamic range of the instrument when scanning using the absorbance blank.

Figure 82 Set Scan Gains Screen



To change the scan gains, follow these steps.

1. Verify that both sources are turned on and that they have stabilized for at least one hour.
2. Place a cuvette containing blank solution in the cell holder.
3. Click on the Set Scan Gains Now button. The instrument adjusts the scan gain values to compensate for the absorbance of the blank.


NOTICE

The instrument will continue to use the new gain values until the gain is adjusted using either the Calibrate Wavelength or the Set Scan Gains function. It may be desirable to repeat the above procedure with nothing in the sample compartment after changing the gain for a significantly absorbing blank.

Reset Source Hours

This function resets the internal usage counter for both, the Visible Source and the UV Source. Click on Reset Hours for Visible Source or Reset Hours for UV Source to reset the counter.

Figure 83 Reset Source Hours Screen



A message will prompt the user to confirm the action or abort.


NOTICE

Do not reset the usage counter unless the respective source has been replaced. Valuable service information will be lost otherwise.

Service Diagnostics

The *Service Diagnostic* function is password protected and will only be accessed by service personnel and manufacturing.

Figure 84 Service Diagnostics Screens



The first window is the default. When clicking on *Change Password*, the window will be enlarged to allow the operator to enter and confirm a new password. Three attempts to enter the correct password are allowed. These passwords are not case-sensitive.

Calibration Info

More information about this function is available in the *DU 800 Service Manual*.


Figure 85 Calibration Information Screen

SCAN GAINS	
Wavelength	Gain
190.0 - 321.0 nm	240
321.1 - 356.0 nm	255
356.1 - 415.0 nm	224
415.1 - 556.0 nm	48
556.1 - 735.0 nm	24
735.1 - 1100.0 nm	40
OTHER	
Grating Constant	1900.17
Sensor Location	3126
Current Drive Location	6904
D2 Reference Location	6793

Wavelength Drive Run

More information about this function is available in the *DU 800 Service Manual*.

Figure 86 Wavelength Drive Run Screen




"*Current Cycle = X*" will be displayed during a run and then replaced by "Finished Run".

DRP Scan

More information about this function is available in the *DU 800 Service Manual*.

Figure 87 DRP Scan Screen



The DRP scan window is very similar to the 'Baseline' or 'Run Baseline Test' window. The *Commands* menu items perform the same functions as the buttons. The *Axis* menu items are described in RediScan.

The default Wavelength parameters are 190.0 nm and 1100.0 nm (range from 190.0 nm to 1100.0 nm). The default Counts parameters are 0 to 50,000 (range is from 0 to 50,000)


The Blank scan is taken automatically before the actual DRP scan.

The scan can be printed but not saved to a file.

Blocked Beam Noise

More information about this function is available in the *DU 800 Service Manual*.


Figure 88 Blocked Beam Noise Screen





Erase EEPROM

More information about this function is available in the *DU 800 Service Manual*.

Figure 89 Erase EEPROM Attention Message Screen



Help Menu



The first three sections of this menu allow the user to access the help file in different ways. The *Contents* item calls the "List of Topics" and lets you go to a specific item systematically. The *Search by Index* item lets you search for a specific item by index.

The *System Software*, *Applications Software*, and *Support and Service* items will bring you right to the top level of the respective section.


The *Installation* and *Getting Started* items should be accessed when installing the system and to introduce a user to the basic functionality, using three examples. The *Accessories* item gets you right to the appropriate accessory section.

The *Adobe Acrobat Files (PDF)* item calls the Adobe Acrobat Reader and automatically loads the respective *Installation and Operating Instructions* or other documentation. The respective PDF files will be copied to the hard disk during the installation process to make the information readily available.

Adobe Acrobat might have been installed during the *DU 800 System and Applications Software* installation. However, the installation of Adobe Acrobat during installation is

optional. If a valid Adobe Acrobat installation cannot be found, an error message appears. In this case, you might want to install or re-install Adobe Acrobat.

Figure 90 About This Program Screen



The *About this Program* item provides program, version and copyright information.

Applications Software


The applications software is described in Help and the optional Software Reference Manual.

Sampling Accessories

Transport

Ambient or water temperature-controlled *Single Cell Holders* can be installed on the *Static Mount* and do not require a transport. However, they may also be mounted on the transport. Multi-Position Cell Holders require a transport.

Figure 91 Transport Accessory



Two types of transports are available:

- **Standard Transport**

For ambient and water temperature-controlled cell holders and all types of cuvettes.

- **High Performance Transport**


Required for Peltier temperature-controlled cell holders. This transport contains a fan to remove excess heat and provides additional accuracy in positioning microcells.

Refer also to the *Installation Instructions, Transport Accessory (P/N 512870)*.

Cell Holders

The installed cell holder must be specified by the user in the Transport/ HOLDERS tab of the *Accessories* window. All available cell holders from Beckman Coulter can be configured as well as a custom cell holder with up to 12 positions.

Figure 92 Cell Holders



The following cell holders are available:

- Single Cell Holders
 - Ambient (adjustable)
 - Water-temperature controlled
 - Peltier-temperature controlled*
- Single Cell Holder, Rectangular, Ambient
- Single Cell Holder, Cylindrical, Ambient
- Turbidity Cell Holder (Single), Ambient
- 6-Position Cell Holder
 - Ambient
 - Water-temperature controlled
 - Peltier-temperature controlled*
- 6-Position Microcell Holder, Ambient
- 7-Position Cell Holder, Ambient
- 8-Position Cell Holder, Ambient
- 12-Position Microcell Holder, Ambient
- T_m Microcell Holder, Peltier-temperature controlled**

* These cell holders require a Peltier Temperature Controller and the High Performance Transport to remove the excess heat.


** This cell holder requires the High-Performance Peltier Temperature Controller and the High Performance Transport to remove the excess heat.

Refer also to the *Installation Instructions, Single Cell Holders (P/N 512874)* and *Installation Instructions, Multicell Holders (P/N 512875)*.

Sipper Accessory

The sipper consists of a peristaltic pump, which is used to aspirate a sample into a flow cell. After the sample is aspirated into the flow cell, a reading is taken automatically.

Figure 93 Sipper Accessory



The *Sipper Installed* item in the Sipper tab of the *Accessories* window must be checked when a sipper is installed. The sipper parameters are in the same tab and may be optimized for the samples.

Refer also to the *Installation Instructions, Sipper Accessory (P/N 512873)*.

Sipper Controls

The sipper panel has four control buttons.

[FILL/BLANK] - Aspirates the sample for the set *Fill Time*, waits the set *Read Delay Time*, and then reads the sample as a blank.

[RETURN] - Returns the sample for the set *Return Time*.

[FLUSH] - Runs the pump in the forward direction as long as the button is pressed. Used to flush the flow cell with solvent, sample, etc. without taking a subsequent measurement. After flushing the flow cell with solution, the user may take a reading as an alternative to [FILL/READ].

[FILL/READ] - Aspirates the sample for the set *Fill Time*, waits the set *Read Delay Time*, and then reads the sample as a sample.

[RETURN], [FILL/BLANK] and [FLUSH] are active at all times. [FILL/READ] is only active when "Sipper" and/or "Batch Sampler" are selected in the *Sampler* tab of the Method window.


Sipper Operation

1. Verify that the desired fill time, return time, read delay time, and sipper speed have been set.
2. Clean the flow cell by flushing with distilled water, then an alkaline solution, distilled water again, and finally the solvent used for the analysis. Several seconds with each solution should be sufficient for cleaning, if the flow cell was cleaned after the previous use.
3. Make sure that *Sipper* is selected in the *Sampler* tab of the Method window.
4. Blank - Aspirate the blank using [FILL/BLANK]. This will blank the system.
5. Flush the solvent from the flow cell using [FLUSH] or return the solvent using [RETURN]. It is important to have at least a 5 second air flush between each sample to minimize carryover.
6. Analyses requiring a re-read of standards only. Aspirate the standard using [FILL/READ]. Then put the focus onto the desired Standard and Replicate by clicking on the respective row in the grid below the standard curve. Flush or return each sample.
7. Samples - If desired, change the default sample identification (Sample n) to a meaningful description. Aspirate and read each sample using [FILL/READ]. Flush or return after each sample.
8. When the analysis is complete, clean the flow cell using solvent, then an alkaline solution, and distilled water. When the flow cell is clean, leave it filled with distilled water. It is important that the flow cell never be left to air dry.
9. At the end of the day, the daily cleaning procedure described in the *Section Cleaning of Flow Cells* in the *Installation Instructions, Single Cell Holders (P/N 512874)* and *Installation Instructions, Multicell Holders (P/N 512875)* should be used. Shortcuts in the cleaning procedure usually prove to be self-defeating and should be avoided since a build up of contaminants may occur. The tubing should remain attached to the flow cell. When use is resumed, the flow cell need only be emptied, flushed several times, then filled with sample.

Batch Sampler Accessory

The batch sampler is an autosampler capable of holding multiple samples in test tubes. The solutions held in the test tubes are aspirated into the sipper flow cell for analysis, using the sipper pump. After the analysis, the samples either are returned to the batch sampler or are sent to waste. If the batch sampler has a wash station, the flow cell can be rinsed between samples.

Figure 94 Batch Sampler Accessroy



The batch sampler offered by Beckman Coulter is capable of holding up to 114 samples. Currently, this is the only batch sampler that is supported by the software.

NOTICE

The sipper accessory is required.

The batch sampler must be selected in the Batch Sampler tab of the *Accessories* window for proper operation. The fill time, return time, read delay time, and the sipper speed parameters in the same tab might be optimized for the samples. The recommended minimum sample volume with the batch sampler is 1.5 mL.

Refer also to the *Installation Instructions, Batch Sampler Accessory (P/N 512871)*.

Beckman Coulter (ISCO) Batch Sampler

Stop Rack - The batch sampler is provided with 18 white test tube racks and 1 red (stop) rack. The stop rack is sensed by the batch sampler. When the last sample position of the stop rack is reached, the analysis will stop, even if there are more samples to read.

Manual Advance Button - This button is located on the back of the batch sampler and will raise the aspirator arm, if it is lowered, and advance the racks one sample position.

Manual Movement of the Aspirator Arm - The aspirator arm can be moved manually up and down, to and from the wash station (if installed), and to the next tube position. To move it, click on the respective button in the Batch Sampler tab.

Automatic Operation

1. If the batch sampler aspirator tube is positioned above the last position in the red stop rack, press the Manual Advance button on the back of the batch sampler to advance the racks one sample position.
2. Clean the flow cell by flushing with distilled water, then an alkaline solution, distilled water again, and finally the solvent used for the analysis. Several seconds with each solution should be sufficient for cleaning, if the flow cell was cleaned after the previous use. Empty the flow cell by pressing [FLUSH] for about 5 seconds.
3. Make sure that *Batch Sampler* is selected in the *Sampler* tab of the Method window.
4. Place all solution in 13 x 100 mm test tubes.
5. Blank - Place the blank directly under the aspirator tube. Click on the *Lower Arm* button to lower the aspirator arm.

Press [FLUSH] until the blank solution is aspirated into the flow cell and blank the system. Then press [FLUSH] to aspirate the solution to waste. Raise the aspirator arm by clicking on the *Raise Arm* button.

6. Analyses with standards only. Place the standard(s) in order, starting with the tube directly under the aspirator tube. Click on the GO icon. The batch sampler will aspirate each standard, read it, return it as specified, and move to the next position. Standard readings are taken until either the last standard is read or the last position of the red stop rack is read.
7. Samples - Place the samples in the tubes, starting with the tube directly under the aspirator tube. Do not leave any empty positions. Click on the START icon. The samples are analyzed. The default sample identifications will be applied (Sample n), where n is a sequential number starting with 1.
8. The analysis will terminate when the sample in the last position of the red stop rack has been read.
9. When the analysis is complete, clean the flow cell using solvent, then an alkaline solution, and distilled water. When the flow cell is clean, leave it filled with distilled water. It is important that the flow cell never be left to air dry.
10. At the end of the day, the daily cleaning procedure described in the *Section Cleaning of Flow Cells* in the *Installation Instructions, Single Cell Holders (P/N 512874)* and *Installation Instructions, Multicell Holders (P/N 512875)* should be used. Shortcuts in the cleaning procedure usually prove to be self-defeating and

should be avoided since a build up of contaminants may occur. The tubing should remain attached to the flow cell. When use is resumed, the flow cell need only be emptied, flushed several times, then filled with sample.

Manual Operation

It is sometimes desirable or necessary to place samples in the batch sampler and manually aspirate them into the instrument for readings, rather than analyzing them automatically.

Examples of situations where this may occur are:

1. Blanking before an automatic analysis.
2. After automatic analysis of standards, when one or more standards need to be re-run.
3. When the batch sampler is set up and used for a routine assay. Between assays, a wavelength scan needs to be taken on a sample.

To perform these analyses:

1. Verify that the desired operating parameters are set. In the third example above, the RediScan mode can be used.
2. Place the sample in the batch sampler under the aspirator arm.
3. Click on the *Lower Arm* button to lower the aspirator arm.
4. Press [FLUSH] until the flow cell is filled to aspirate the solution manually.
5. Blank or take a reading.
6. After the reading is taken, press [FLUSH] to aspirate the solution to waste.
7. To read another solution, located in the next position in the batch sampler, click on the appropriate button to raise the aspirator arm, advance to the next sample, and lower the arm again. Repeat the directions in steps 4 to 6 to take another reading.


Peltier Temperature Controller

NOTICE

To control the sample temperature with the Peltier Temperature Controller, the High Performance Transport is required to remove the excess heat from the cell holder.

The Peltier method of temperature control allows both heating and cooling of the sample, hence the ability to cool below ambient temperature. If use of the Peltier Temperature Controller is desired, make sure the *Temperature Controller Enabled* box in the Temp Controller tab of the *Accessories* window is checked.

Figure 95 Peltier Temperature Controller



Please notice that the glass (silica or quartz) of the cuvettes provides insulation between the sample contained in the cell and the cell holder, which is heating/cooling the sample. A moderate amount of time (~3 minutes) is required to heat the sample after the cell holder has equilibrated at the desired temperature.

The temperature controller is turned on using the switch on the back. When turned on, the POWER indicator light on the front of the controller is illuminated continuously.

When the temperature controller is activated, the OPERATING indicator light on the front of the controller blinks and the cooling fan, located under the transport, turns on (High Performance Transport only). The fan remains on for about 5 minutes after the temperature controller is disabled to remove the remaining heat from the cell holder.

Peltier temperature-controlled cell holders are plugged into the temperature connector on the left sample compartment wall.

Refer also to the *Installation Instructions, Peltier Temperature Controller (P/N 512872)*.

Standard Controller


The standard Peltier Temperature Controller is an accessory used with the temperature-controlled versions of the *Single Cell* and *Multicell Holders* to control the temperature of the sample(s) in the range of 20 to 40 °C.

The standard Peltier Temperature Controller is not compatible with the Experimental T_m Analysis application.

High Performance Controller

The High Performance Peltier Temperature Controller may be used instead of the standard controller for all cell holders.

The *Experimental T_m Analysis* application requires the High Performance Temperature Controller and the *T_m Microcell Holder*. This combination allows precise temperature control and fast ramping in the range of 0 to 110 °C. The temperature settings are done in the *Experimental T_m Analysis* program instead of the *Temp Controller* tab.



Technical Specifications

Storage and Transport


The power should be turned off on the DU 800 Spectrophotometer when not in use for several days or more. The sample compartment cover should be closed. It is necessary that the instrument remain in the upright position, on a flat surface, so that the optics will remain stable.

The instrument should be transported in the upright position, with care taken not to jolt, bounce, or shake the instrument during transport. Pack the instrument in the original shipping container, if still available.

Sample Compartment Configuration

The following diagram shows the location and size of the beam in the sample compartment. Notice that the beam focus is near the right hand side of the sample compartment. Sampling accessories are installed in this location and will position the sample at the beam focus.

Figure 96 Diagram of Location and Size of the Beam in Sample Compartment



Performance Specifications

Scan Speeds	120, 240, 600, 1200, 2400 nm/min
Data Collection Rate	20 Samplings per Second
Response Time	0.05 Seconds
Wavelength Range	190 to 1100 nm
Wavelength Setability	0.1 nm Increments
Wavelength Accuracy - At 656.1 nm - Full Range	±0.2 nm ±0.5 nm
Wavelength Repeatability - At 656.1 nm - Full Range	±0.1 nm ±0.2 nm
Spectral Bandwidth (from 200 to 680 nm)	<=1.8 nm
Photometric Readout	-0.300 to 3.000A or 0.0 to 200%T
Photometric Accuracy (at 1A with NIST 930D solid filter at 546 nm)	±0.005A
RMS Noise (at 0A, average of 10 standard deviations of 10 readings at 0.05 sec intervals, at 500 nm)	<0.0002A RMS
Stray Light (measured using NaI at 220nm, per ASTM E387-84)	<0.05%
Stability (0A, constant ambient conditions, measured for one hour at 340 nm after one hour warmup)	<0.003A/hr
RMS Baseline Flatness (from 200 to 900 nm, at 0 A)	±0.001A RMS

Physical & Environmental Specifications

Width	69 cm (27 inches)
Depth	53 cm (21 inches)
Height	26 cm (10 inches)
Weight	33 kg (73 pounds)
Line Voltage	100-240V ±10%
Frequency	50/60 Hz
Power	200 watts typical
Ambient Temperature Operating Range	+15 to 40°C (59 to 104°F)
Humidity	<85% maximum relative humidity, not to exceed 32.5°C WBT

These specifications apply to the optical bench, which is the DU 800 Spectrophotometer without a computer.

Maintenance

NOTICE

Refer to the SAFETY NOTICE before following replacing any items. Refer all servicing for procedures not contained in this section to qualified service personnel.

General Information

The DU 800 UV/Visible Spectrophotometer is a PC controlled, optical instrument. For the system to operate properly, it is necessary to have adequate ventilation, as described in Installation. It is necessary to operate the instrument below the specified maximum temperature of 40°C (104°F) for adequate cooling.

For the optical components to maintain their quality, it is necessary to operate the instrument in a dust-free, non-corrosive environment. If it is required that the instrument be used in a dusty or corrosive environment, the optical surfaces must be cleaned or replaced periodically. Contact Beckman Coulter Service for more information.

One component of the instrument that can be user-serviced to prevent an interruption in operation is the UV source lamp. With time, its output gradually decreases, close to one-half of the original value, in approximately 1000 hours. At this point in time, it is appropriate to replace the UV source, even though it is not burned out. The length of time that the UV source has been illuminated is shown in the System Status window, accessible in the Diagnostics menu.


A burned out UV source is indicated by the message: "UV LAMP Failure". Instructions for replacing the UV source can be found in UV Source Replacement.

If decreased sensitivity is seen in the visible range, the visible source should be examined to see if a coating has built up on the inside of the glass envelope. If so, the lamp should be replaced. Instructions for replacing the visible source can be found in Visible Source Replacement.

System Status

The System Status window, accessible in the Diagnostic Menu, should be examined as part of the Preventive Maintenance program.

Figure 97 System Status Window Screen




The window shows, among other things, the amount of time that both source lamps have been in use. The UV source should be changed at 1000 hours. The visible source should be changed at 2000 hours.

Fuse Replacement

The only user accessible fuse is the power fuse, located above the receptacle where the power cord is plugged into the instrument. The instrument is shipped with the fuse installed.

1. Turn the instrument OFF.
2. Turn the instrument around so that the back can be accessed.
3. Unplug the power cord from the instrument.
4. The power line module contains a power on/off switch, a power cable receptacle, and a fuse compartment with one fuse installed. To access the fuse, pull and lower the fuse compartment as shown below. The fuse is specified for 100 to 240 VAC (T3A 250V), P/N 890895.

Figure 98 Fuse Compartment



WARNING

For continued protection against risk of fire, replace the fuses only with the type and current rating specified above.

ATTENTION

Afin d'assurer une protection permanente contre les risques d'incendie, remplacer uniquement par un fusible de même type et valeur.

5. Insert the fuse compartment and press to secure.
6. Plug the power cord back into the instrument.
7. Turn the instrument around so that it is facing forward.
8. Turn the instrument ON.

UV Source Replacement

Parts Required: UV Source, P/N 514366

1. Turn the instrument off, unplug the power cord and allow the instrument to cool down for 15 minutes.

CAUTION

To avoid risk of electrical shock, disconnect power cord to instrument before changing either source.

Attention


Risque de choc électrique. Débrancher l'appareil avant de changer de source.

The UV lamp operates at a high temperature. To prevent burns, allow at least 15 minutes for the instrument and sources to cool before handling internal components.

The UV lamp generates UV light. Do not look directly at an operating lamp without wearing proper UV protective eye glasses.


2. Locate the source access door on the top left-hand of the instrument cover, then loosen the thumb screw and open it.

Figure 99 Location of Source Access Door on DU 800




3. To unplug the connector for the UV source squeeze on the tabs on the sides of the connector, then pull up to release.

Figure 100 Unplugging UV Source Connector




4. Unscrew the two thumb screws that hold the UV source in position.

Figure 101 Unscrewing Thumb Screws securing UV Source



5. Remove the source by lifting it straight up.

Figure 102 Removing UV Source




6. Unpack the new source, being careful not to touch the glass envelope on the source. Carefully push the new source straight down, until it is seated flat against the source housing. Secure by tightening the two thumbscrews and then plug in the connector.

NOTICE

Do not touch the glass envelope on the new source. If it is touched, clean with alcohol and a soft cloth.

7. Close the source access door and tighten the thumbscrew.
8. Plug in the instrument power cord and turn the spectrophotometer on. Then start the DU 800 System and Applications Software, wait until the system has re-initialized, and press the Continue button to enter the software.
9. Click on Reset Source Hours in the Diagnostic menu under User Diagnostics and reset the hours for the UV source.

Figure 103 Reset Source Hours Screen

**Visible Source Replacement****Part I. Replacing the Source**

Parts required: Visible Source, P/N514259

1. Unplug the instrument power cord. Allow the instrument to cool for 15 minutes.

CAUTION

To avoid risk of electrical shock, disconnect power cord to instrument before changing either source.


Attention

Resque de choc électrique. Debrancher l'appareil avant de changer de source.

The visible lamp operates at a high temperature. To prevent burns, allow at least 15 minutes for the instrument and sources to cool before handling internal components.


2. Locate the source access door on the top left-hand of the instrument cover, then loosen the thumb screw and open it.

Figure 104 Location of Source Access Door on DU 800




3. Unscrew the two thumbscrews that hold the visible source mounting bracket in position.

Figure 105 Unscrewing Thumbscrews Holding Visible Source



4. Remove the source mounting bracket by lifting straight up. Rotate the source mounting bracket so that the visible source is accessible.

Figure 106 Removing Visible Source




5. Carefully remove the old source by pulling it straight out. Unpackage the new source, being careful not to touch the glass envelope on the source. Hold the source with a tissue or Kim-Wipe while inserting it into the source mount in the same location as the old source.

NOTICE

Do not touch the glass envelope on the new source. If it is touched, clean with alcohol and a soft cloth.

Figure 107 Installing New Visible Source



6. Replace the source mounting bracket and tighten the two thumbscrews.


Part II. Peaking the Source

Tool required: Peaking tool (provided in the shipping kit)

1. Make sure the instrument is turned on and has been initialized by the DU 800 System and Applications Software.
2. Turn the Visible Source ON.
3. Select the Kinetics/Time application and choose the <Default Method>.
4. Click the Edit Method button and change the Interval time to 1.5 seconds, then click OK.
5. Click on the Blank icon button to blank the instrument on air and then click on the Read button. A reading of approximately 0 Abs. is displayed as data is collected.
6. To peak the source, slightly loosen the two thumbscrews that secure the visible source mounting bracket.


7. Insert the peaking tool into the adjustment hole on the front corner of the source mounting bracket.

Figure 108 Using Peaking Tool



8. Turn the peaking tool a small amount in the clockwise direction and observe the readings. The goal is to achieve the lowest reading possible.
 - a. If the readings are lower than 0 Abs., continue turning clockwise until the lowest reading is obtained.
 - b. If the readings are above 0 Abs., turn counter-clockwise until the lowest reading is obtained.
 - c. If the readings reach -0.3 Abs., restart the run with a new blank reading.
 - d. Repeat the above steps until the lowest reading is obtained.
9. Tighten the two thumbscrews that secure the visible source mounting bracket. See Figure 2.
10. Close the source access door and tighten the thumbscrew.
11. Click on Reset Source Hours in the Diagnostic menu under User Diagnostics and reset the hours for the visible source.

Figure 109 Reset Source Hours Screen




Troubleshooting

Startup Diagnostic Tests

When the *DU 800 System and Applications Software* is launched and an instrument is detected on the USB port, the software downloads the firmware to the DU 800, builds the wavelength calibration table, and then executes a series of diagnostic tests to ensure proper operation of major system components. As the tests are performed, the entire progress and the individual results are displayed. When the tests are complete, the software may be used.

Figure 110 DU 800 System Initialization Screen



Following the test name is Passed or >>> FAILED <<<. If any of the tests fail, click on the *Cancel* button and then power down the instrument. Wait a few seconds and then power it up again. If the failure persists, the instrument may be partially operational as described in the following table.

The *Visible Lamp* and *Light Path* failures can be corrected by the user. If the problem persists, contact the local Beckman Coulter service office. If any other test fails, contact the local Beckman Coulter office.

Failed Test	Explanation
Detector	Readings cannot be taken. The instrument can be used for other functions.
Gain	Readings cannot be taken. The instrument can be used for other functions.
Visible Lamp	Replace the visible source. Directions are provided in Visible Source Replacement.
Light Path	Something is blocking the light path through the sample compartment. Remove the sampling accessory and re-initialize the system.
Shutter	Readings cannot be taken. The instrument can be used for other functions.
Filter	Readings cannot be taken. The instrument can be used for other functions.
EEPROM	Readings cannot be taken. The instrument can be used for other functions.
Wavelength Drive	The wavelength may not be accurate. Use the instrument only if the accuracy of the readings can be verified. Contact the Beckman Coulter service.

Operational Failures

There are three failures that can occur during instrument operation. These are displayed in the Permanent Menu Bar on the left-hand side. The failure is displayed until the condition is corrected.

Message	Recommended Action
VISIBLE LAMP Failure	Replace the visible source. Directions are given in Visible Source Replacement.
UV LAMP Failure	Replace the UV source. Directions are given in UV Source Replacement.
Visible Lamp	Replace the visible source. Directions are provided in Visible Source Replacement.
Light Path	Something is blocking the light path through the sample compartment. Remove the sampling accessory and re-initialize the system.
Filter Readings cannot be taken	The instrument can be used for other functions.
Lamp Selector	Readings cannot be taken. The instrument can be used for other functions.
Wavelength Drive	The wavelength may not be accurate. Use the instrument only if the accuracy of the readings can be verified. Contact the Beckman Coulter service.

Operational Messages

The following messages can occur while operating the instrument and are displayed in an error window. In most cases, these errors do not indicate an instrument malfunction but a situation where an action was initiated that the instrument cannot perform.

If the instrument requires operator action to continue, the message is given in the form of a message window and the user must click on OK to remove the window.

Message	Explanation and Recommended Action
Blank Required	Place a cuvette with solvent in the cell holder and click the BLANK icon button. Any reading taken without a blank is invalid.
UV Lamp Required	Turn on the UV source. Wait for the UV source to warm up and blank the instrument. Any reading taken without the proper source on is invalid.
Visible Lamp Required	Turn on the visible source and blank the instrument. Any reading taken without the proper source on is invalid.

DU 800 Spectrophotometer
Installation & Operating Instructions

Service and Support

Please make sure to complete and return the Warranty Registration after installation.

This will guarantee that your system will be registered and that the one-year warranty is applied properly. It also assures quick and easy access to the Beckman Coulter Product Support team for answers to your questions.

Customer Support **(800) 551-1150**

Marketing Support **(800) 742-2345**

For North America Only

Warranty Registration

The Warranty Registration form is included under separate cover. If you have not done so, please complete the form and return to the address indicated.

Beckman Coulter Offices

North America

United States

Beckman Coulter, Inc. (Headquarter)
4300 N. Harbor Boulevard
P. O. Box 3100
Fullerton, California 92834-3100
USA

Phone: General (714) 871-4848
Phone: Sales (800) 742-2345
Phone: Service (800) 551-1150
FAX: (800) 643-4366
FAX: (714) 773-8283

Canada

Beckman Coulter Canada Inc.
6755 Mississauga Road, Suite 600
Mississauga, Ontario, Canada L5N 7Y2

Phone: (800) 387-6799 BCO
FAX: (905) 819-1485

Latin America

Mexico

Beckman Coulter de Mexico
Adolfo Prieto 1213
Colonia del Valle A.P. 44-100
Mexico D.F. 03100

Phone: (52-5) 559-1635
FAX: (52-5) 559-3941
E-Mail: jcamacho@beckman.com

Europe

France

Beckman Coulter France S.A.
Paris Nord II
33 rue des Vanesses
BP 50359
VILLEPINTE
95942 ROISSY CDG Cedex

Phone: 33 (0)1 49 90 90 00
FAX: 33 (0)1 49 90 90 10
E-Mail: bioresearch.france@beckman.com

Germany

Beckman Coulter GmbH
Siemensstrasse 1
D-85716 Unterschleissheim-Lohhof
Phone: 49-89-35-870-0 (Switchboard)
FAX: 49-89-35-870-490 (Gen'l Fax)

Italy

Beckman Coulter S.p.A.
Centro Direzionale Lombardo
Palazzo F/1, Via Roma 108
20060 Cassina De'Pecchi
Milano
Phone: (34) 02 953921
FAX: (34) 02 95392264
E-Mail: beckman_italia@beckman.com

Netherlands

Beckman Coulter (Nederland) B.V.
Nijverheidsweg 21
3641 RP Mijdrecht
P.O. Box 47
3640 AA Mijdrecht,
Phone: 31-(0)297-230630
FAX: 31-(0)297-288082

Spain

Beckman Coulter España S.A.
Avda. del Llano Castellano 15
28034 Madrid

Phone: (34) 91 383 6080
FAX: (34) 91 383 6096
E-Mail: bioresearch.espana@beckman.com

Sweden

Beckman Coulter AB
Archimedesvagen 7
Box 11156
S-16811 Bromma

Phone: 46 8 564 859 00
FAX: 46 8 564 859 01
E-Mail: bc@beckmancoulter.se

Switzerland

Beckman Coulter Intl SA (Swiss Domestic Operation)
22, Rue Juste-Olivier
Case Postale 301-303
CH-1260 Nyon , Switzerland

Phone: (41) 22 994 0707
FAX: (41) 22 994 0700

Turkey

Beckman Coulter Ltd/Turkey
E-5 Yanyol Fatih Cad.
81410 Soganlik Kartal
Istanbul

Phone: + 90 216 309 1900
FAX: + 90 216 309 0090

United Kingdom

Beckman Coulter (U.K.) Limited
Oakley Court
Kingsmead Business Park
London Road
High Wycombe
Buckinghamshire HP11 1JU, England

Phone: (44) 1494-441181
FAX: (44) 1494-447558
E-Mail: beckmancoulter_uk@beckman.com

Asia/Pacific

Australia

Beckman Coulter Australia Pty. Ltd.
24 College Street
Gladesville, NSW, 2111

Phone: (61) 2-9844-6000

Phone: 1 800 060 880

FAX: (61) 2-9844-6096 Sales & Marketing

FAX: (61) 2-9844-6098 Service

China

Beckman Coulter Inc. Beijing Representative Office
Unit 2005A, 2006-2009, East Ocean Center
24 Jian Guo Men Wai Avenue
Beijing 100004

Phone: 86-10-6515-6028

FAX: 86-10-6515-6025

Hong Kong

Beckman Coulter Hong Kong Ltd.
12th Floor, Oxford House
979 King's Road
TaiKoo Place

Phone: 852-2814-7431

Phone: 852-2814-0481

FAX: 852-2814-1599

FAX: 852-2873-4511

E-Mail: beckmancoulter_hk@beckman.com

Japan

Beckman Coulter K.K.
Toranomom 37 Mori Building
3-5-1 Toranomom, Minato-ku
Tokyo 105-0001

Phone: +81 (3) 5404-8359

FAX: +81 (3) 5404-8436

Singapore

Beckman Coulter Singapore Pte Ltd.
116 Changi Road
Unit 03-01/02
Singapore 419718
Phone: (65) 339 3633
FAX: (65) 336 6303

Taiwan

Beckman Coulter Taiwan Inc., Taiwan Branch
8th Floor
216 Tun Hwa South Road, Section 2
Taipei 106
Taiwan
Phone: 886-2-2378-3456
Phone: 886-2-2378-8000
FAX: 886-2-2377-0408

Eastern Europe, Middle East, Africa

Beckman Coulter Intl SA
22, rue Juste-Oliver
P.O. Box 303
1260 Nyon, Switzerland
Phone: (41) 22 994 0707
FAX: (41) 22 994 0700

South Africa

Beckman Instruments Pty. Ltd.(Johannesburg)
(Postal) P.O. Box 1949
Halway House 1949
(Physical) Stand 1A Fedsure Park
Tonetti Street, 1685 Halway House
Johannesburg
Phone: (27) 11 805-2014 or (27) 11 805-1015
FAX: (27) 11 805-4120